



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121625

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/22/2022	11:00
2	CAL2	1	Water	NA	NA	20.7	7.00	08/22/2022	11:01
3	CAL3	1	Water	NA	NA	20.7	10.03	08/22/2022	11:02
4	ICV	1	Water	NA	NA	20.8	6.98	08/22/2022	11:05
5	CCV1	1	Water	NA	NA	20.7	2.01	08/22/2022	11:06
6	N4279-02	1	Water	NA	NA	20.3	6.35	08/22/2022	11:10
7	N4279-02DUP	1	Water	NA	NA	20.7	6.36	08/22/2022	11:11
8	CCV2	1	Water	NA	NA	20.8	12.03	08/22/2022	11:15

WB 121625

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N4279

WorkList ID : 162486

Department : Wet-Chemistry

Date : 08-22-2022 10:40:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/19/2022	Water	N4279-02	pH	Cool 4 deg C	TETR06	K11	TT-050-IDWGW-20220819	08/19/2022	9040C

Date/Time 08/22/22 10:42
Raw Sample Received by: JBLWCI
Raw Sample Relinquished by: CL SM

Date/Time 08/22/22 12:00
Raw Sample Received by: CL SM
Raw Sample Relinquished by: PCAC