

LB121626


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57466



A= 0.0000e+000

B= 1.7225e-004

C= 1.4964e-002

Rho= 0.9998466

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	0.010	0.010	-30	0.000	-30					0.0
0.2	0.200	0.182	-0.018	967	0.0 %	967					-9
2.5	2.500	2.435	-0.065	14050	0.0 %	14050					-3
5.0	5.000	5.078	0.078	29393	0.0 %	29393					2
7.5	7.500	7.583	0.083	43935	0.0 %	43935					1
10.0	10.000	9.913	-0.087	57466	0.0 %	57466					-1

LB121626 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Aug 2022 10:59:04

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		-30	-	0.0000	7471B	PPB	22 Aug 2022 11:00:17	S	Std
0.2 - 1		967	-	0.2000	7471B	PPB	22 Aug 2022 11:02:33	S	Std
2.5 - 1		14050	-	2.5000	7471B	PPB	22 Aug 2022 11:04:51	S	Std
5.0 - 1		29393	-	5.0000	7471B	PPB	22 Aug 2022 11:07:08	S	Std
7.5 - 1		43935	-	7.5000	7471B	PPB	22 Aug 2022 11:09:24	S	Std
10.0 - 1		57466	-	10.0000	7471B	PPB	22 Aug 2022 11:11:41	S	Std
ICV97 - 1	ICV97	21344	3.6914	-	7471B	PPB	22 Aug 2022 11:14:36	U	SMPL
ICB97 - 1	ICB97	-118	-0.0054	-	7471B	PPB	22 Aug 2022 11:16:50	U	SMPL
CCV79 - 1	CCV79	29443	5.0864	-	7471B	PPB	22 Aug 2022 11:19:07	U	SMPL
CCB79 - 1	CCB79	-130	-0.0074	-	7471B	PPB	22 Aug 2022 11:21:22	U	SMPL
CRA - 1	CRA	1008	0.1886	-	7471B	PPB	22 Aug 2022 11:23:39	U	SMPL
HighStd - 1	HighStd	58038	10.0118	-	7471B	PPB	22 Aug 2022 11:25:54	U	SMPL
ChkStd - 1	ChkStd	41567	7.1747	-	7471B	PPB	22 Aug 2022 11:28:11	U	SMPL
PB147169BL - 1	PB147169BL	-191	-0.0179	-	7471B	PPB	22 Aug 2022 11:30:30	U	SMPL
PB147169BS - 1	PB147169BS	20293	3.5104	-	7471B	PPB	22 Aug 2022 11:32:47	U	SMPL
N4255-01 - 1	POLYLINER-COMP	-80	0.0012	-	7471B	PPB	22 Aug 2022 11:35:04	U	SMPL
N4257-01 - 1	VNJ-210	13401	2.3232	-	7471B	PPB	22 Aug 2022 11:37:22	U	SMPL
N4258-01 - 1	RBR-21153	2246	0.4018	-	7471B	PPB	22 Aug 2022 11:39:39	U	SMPL
N4258-03 - 1	280251	2223	0.3979	-	7471B	PPB	22 Aug 2022 11:41:57	U	SMPL
N4261-01 - 1	C5	3939	0.6934	-	7471B	PPB	22 Aug 2022 11:44:12	U	SMPL
CCV80 - 1	CCV80	28690	4.9567	-	7471B	PPB	22 Aug 2022 11:46:27	U	SMPL
CCB80 - 1	CCB80	-156	-0.0119	-	7471B	PPB	22 Aug 2022 11:48:43	U	SMPL
N4264-01 - 1	B-3	6971	1.2157	-	7471B	PPB	22 Aug 2022 11:55:35	U	SMPL
N4264-01DUP - 1	B-3DUP	6841	1.1933	-	7471B	PPB	22 Aug 2022 11:57:50	U	SMPL
N4264-01MS - 1	B-3MS	25999	4.4932	-	7471B	PPB	22 Aug 2022 12:00:06	U	SMPL
N4264-01MSD - 1	B-3MSD	26512	4.5816	-	7471B	PPB	22 Aug 2022 12:02:22	U	SMPL
N4264-01LX5 - 1		1269	0.2335	-	7471B	PPB	22 Aug 2022 12:04:39	U	SMPL
N4264-01A - 1		31916	5.5124	-	7471B	PPB	22 Aug 2022 12:06:56	U	SMPL
N3980-01 - 1	LOD-MDL-SOIL-01-QT3-2022	154	0.0415	-	7471B	PPB	22 Aug 2022 12:11:48	U	SMPL
N3980-02 - 1	LOQ-SOIL-02-QT3-2022	1021	0.1908	-	7471B	PPB	22 Aug 2022 12:14:05	U	SMPL
CCV81 - 1	CCV81	29448	5.0873	-	7471B	PPB	22 Aug 2022 12:16:22	U	SMPL
CCB81 - 1	CCB81	-143	-0.0097	-	7471B	PPB	22 Aug 2022 12:18:38	U	SMPL