



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/23/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 08/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 08:05
Out Date: 08/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121628

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4282-01	EXXS4	1	0.91	8.86	9.77	8.58	86.6	
N4282-02	EXXS5	2	0.93	8.82	9.75	8.54	86.3	
N4282-03	EXXS6	3	0.93	8.73	9.66	8.26	84.0	
N4282-04	EXXS6D	4	0.93	8.73	9.66	8.26	84.0	
N4282-05	EXXS6S	5	0.93	8.73	9.66	8.26	84.0	
N4282-06	EXXS7	6	0.91	8.87	9.78	8.35	83.9	
N4282-07	EXXS8	7	0.88	8.95	9.83	8.34	83.4	
N4282-08	EXXS9	8	0.93	8.96	9.89	8.46	84.0	
N4282-09	EXXT0	9	0.91	8.82	9.73	8.33	84.1	
N4282-10	EXXT1	10	0.92	8.62	9.54	7.88	80.7	
N4282-11	EXXT2	11	0.98	8.68	9.66	8.45	86.1	
N4282-12	EXXT3	12	0.91	9.06	9.97	8.67	85.7	
N4282-13	EXXT4	13	0.88	8.73	9.61	8.24	84.3	
N4282-14	EXXT5	14	0.82	9.07	9.89	8.69	86.8	
N4282-15	EXXT6	15	0.93	8.89	9.82	8.47	84.8	
N4282-16	EXXT7	16	0.91	8.77	9.68	8.46	86.1	
N4282-17	EXXT8	17	0.93	8.84	9.77	8.57	86.4	
N4282-18	EXXT9	18	0.91	9.07	9.98	8.59	84.7	
N4282-19	EXXW0	19	0.91	8.84	9.75	8.29	83.5	
N4282-20	EXXW1	20	0.88	8.98	9.86	8.65	86.5	
N4282-21	EXXW2	21	0.94	8.77	9.71	8.56	86.9	
N4282-22	EXXW3	22	0.91	8.62	9.53	8.21	84.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$