


LB121634

7470A

INSTRUMENT ID: LVD

Linear



A= 0.0000e+000

B= 2.1698e-004

C= 6.9248e-003

Rho= 0.9996445

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.003	-0.003	-44	0.000	-44					-
0.2	0.200	0.187	-0.013	828	0.0 %	828					-7
2.5	2.500	2.447	-0.053	11247	0.0 %	11247					-2
5.0	5.000	5.190	0.190	23889	0.0 %	23889					4
7.5	7.500	7.363	-0.137	33904	0.0 %	33904					-2
10.0	10.000	10.015	0.015	46123	0.0 %	46123					0

LB121634 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Aug 2022 14:48:39

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-44	-	0.0000	7470A	PPB	22 Aug 2022 14:49:37	S	Std
0.2 - 1	50.2	828	-	0.2000	7470A	PPB	22 Aug 2022 14:51:54	S	Std
2.5 - 1	52.5	11247	-	2.5000	7470A	PPB	22 Aug 2022 14:54:11	S	Std
5.0 - 1	55	23889	-	5.0000	7470A	PPB	22 Aug 2022 14:56:28	S	Std
7.5 - 1	57.5	33904	-	7.5000	7470A	PPB	22 Aug 2022 14:58:44	S	Std
10.0 - 1	50	46123	-	10.0000	7470A	PPB	22 Aug 2022 15:01:01	S	Std
ICV98 - 1	ICV98	16718	3.6344	-	7470A	PPB	22 Aug 2022 15:03:48	U	SMPL
ICB98 - 1	ICB98	-154	-0.0265	-	7470A	PPB	22 Aug 2022 15:06:03	U	SMPL
CCV82 - 1	CCV82	22319	4.8498	-	7470A	PPB	22 Aug 2022 15:08:18	U	SMPL
CCB82 - 1	CCB82	-134	-0.0222	-	7470A	PPB	22 Aug 2022 15:10:33	U	SMPL
CRA - 1	CRA	828	0.1866	-	7470A	PPB	22 Aug 2022 15:12:49	U	SMPL
HighStd - 1	HighStd	46597	10.1177	-	7470A	PPB	22 Aug 2022 15:15:04	U	SMPL
ChkStd - 1	ChkStd	30400	6.6032	-	7470A	PPB	22 Aug 2022 15:17:20	U	SMPL
PB147139BL - 1	PBW	-182	-0.0326	-	7470A	PPB	22 Aug 2022 15:19:37	U	SMPL
PB147139BS - 1	LCSW	17856	3.8814	-	7470A	PPB	22 Aug 2022 15:27:29	U	SMPL
N4247-01 - 1	SB-1	-105	-0.0159	-	7470A	PPB	22 Aug 2022 15:29:45	U	SMPL
N4247-02 - 1	SB-2	-42	-0.0022	-	7470A	PPB	22 Aug 2022 15:32:02	U	SMPL
N4247-03 - 1	SB-3	-29	0.0006	-	7470A	PPB	22 Aug 2022 15:34:19	U	SMPL
N4247-04 - 1	SB-4	-10	0.0048	-	7470A	PPB	22 Aug 2022 15:36:36	U	SMPL
N4247-05 - 1	SB-5	-13	0.0041	-	7470A	PPB	22 Aug 2022 15:38:51	U	SMPL
CCV83 - 1	CCV83	21966	4.7732	-	7470A	PPB	22 Aug 2022 15:41:06	U	SMPL
CCB83 - 1	CCB83	-134	-0.0222	-	7470A	PPB	22 Aug 2022 15:43:22	U	SMPL
N4247-06 - 1	SB-6	-5	0.0058	-	7470A	PPB	22 Aug 2022 15:45:38	U	SMPL
N4247-07 - 1	SB-7	-18	0.0030	-	7470A	PPB	22 Aug 2022 15:47:53	U	SMPL
N4257-03 - 1	VNU-210	-46	-0.0031	-	7470A	PPB	22 Aug 2022 15:50:08	U	SMPL
N4259-01 - 1	8218	-15	0.0037	-	7470A	PPB	22 Aug 2022 15:52:24	U	SMPL
N4259-02 - 1	8220	-50	-0.0039	-	7470A	PPB	22 Aug 2022 15:54:40	U	SMPL
N4259-03 - 1	8186-8185-8216-8215	-40	-0.0018	-	7470A	PPB	22 Aug 2022 15:56:57	U	SMPL
N4261-02 - 1	C5	-41	-0.0020	-	7470A	PPB	22 Aug 2022 15:59:13	U	SMPL
N4261-02DUP - 1	C5DUP	-45	-0.0028	-	7470A	PPB	22 Aug 2022 16:01:30	U	SMPL
N4261-02MS - 1	C5MS	17577	3.8208	-	7470A	PPB	22 Aug 2022 16:05:50	U	SMPL
N4261-02MSD - 1	C5MSD	16851	3.6633	-	7470A	PPB	22 Aug 2022 16:08:06	U	SMPL
CCV84 - 1	CCV84	22636	4.9185	-	7470A	PPB	22 Aug 2022 16:10:22	U	SMPL
CCB84 - 1	CCB84	-140	-0.0235	-	7470A	PPB	22 Aug 2022 16:12:38	U	SMPL
PB147140BL - 1	PBW	-63	-0.0067	-	7470A	PPB	22 Aug 2022 16:14:55	U	SMPL
PB147140BS - 1	LCSW	16979	3.6911	-	7470A	PPB	22 Aug 2022 16:17:10	U	SMPL
N4259-04 - 1	8221	-82	-0.0109	-	7470A	PPB	22 Aug 2022 16:19:25	U	SMPL
N4259-04DUP - 1	8221DUP	-14	0.0039	-	7470A	PPB	22 Aug 2022 16:21:42	U	SMPL
N4259-04MS - 1	8221MS	17195	3.7379	-	7470A	PPB	22 Aug 2022 16:23:58	U	SMPL
N4259-04MSD - 1	8221MSD	17471	3.7978	-	7470A	PPB	22 Aug 2022 16:26:14	U	SMPL
N4259-05 - 1	8188	-132	-0.0217	-	7470A	PPB	22 Aug 2022 16:28:32	U	SMPL
N4259-06 - 1	8217	110	0.0308	-	7470A	PPB	22 Aug 2022 16:30:49	U	SMPL
N4259-07 - 1	8223	121	0.0332	-	7470A	PPB	22 Aug 2022 16:33:06	U	SMPL
N4259-08 - 1	8184	-46	-0.0031	-	7470A	PPB	22 Aug 2022 16:35:21	U	SMPL
CCV85 - 1	CCV85	23558	5.1186	-	7470A	PPB	22 Aug 2022 16:37:37	U	SMPL
CCB85 - 1	CCB85	-175	-0.0310	-	7470A	PPB	22 Aug 2022 16:39:52	U	SMPL
PB147186BL - 1	PBW	-129	-0.0211	-	7470A	PPB	22 Aug 2022 16:42:08	U	SMPL
PB147186BS - 1	LCSW	17106	3.7186	-	7470A	PPB	22 Aug 2022 16:44:24	U	SMPL
N3980-07 - 1	LOD-MDL-WATER-01-QT3-2022	131	0.0353	-	7470A	PPB	22 Aug 2022 16:51:48	U	SMPL
N3980-08 - 1	LOQ-WATER-02-QT3-2022	939	0.2107	-	7470A	PPB	22 Aug 2022 16:54:03	U	SMPL
N4279-02 - 1	TT-050-IDWGW-20220819	-95	-0.0137	-	7470A	PPB	22 Aug 2022 16:56:19	U	SMPL
N4279-02DUP - 1	TT-050-IDWGW-20220819DUP	-74	-0.0091	-	7470A	PPB	22 Aug 2022 16:58:35	U	SMPL
N4279-02MS - 1	TT-050-IDWGW-20220819MS	15725	3.4190	-	7470A	PPB	22 Aug 2022 17:00:52	U	SMPL
N4279-02MSD - 1	TT-050-IDWGW-20220819MSD	15396	3.3476	-	7470A	PPB	22 Aug 2022 17:03:09	U	SMPL
CCV86 - 1	CCV86	22614	4.9138	-	7470A	PPB	22 Aug 2022 17:05:26	U	SMPL
CCB86 - 1	CCB86	-182	-0.0326	-	7470A	PPB	22 Aug 2022 17:07:42	U	SMPL
PB147118TB - 1	PB147118TB	-74	-0.0091	-	7470A	PPB	22 Aug 2022 17:09:58	U	SMPL
PB147123TB - 1	PB147123TB	-81	-0.0107	-	7470A	PPB	22 Aug 2022 17:12:13	U	SMPL
N4261-02LX5 - 1		-114	-0.0178	-	7470A	PPB	22 Aug 2022 17:14:29	U	SMPL
N4261-02A - 1		16935	3.6815	-	7470A	PPB	22 Aug 2022 17:16:45	U	SMPL
N4259-04LX5 - 1		-208	-0.0382	-	7470A	PPB	22 Aug 2022 17:19:00	U	SMPL
N4259-04A - 1		16025	3.4841	-	7470A	PPB	22 Aug 2022 17:21:16	U	SMPL
N4279-02LX5 - 1		-211	-0.0389	-	7470A	PPB	22 Aug 2022 17:23:32	U	SMPL
N4279-02A - 1		17265	3.7531	-	7470A	PPB	22 Aug 2022 17:25:48	U	SMPL
CCV87 - 1	CCV87	23696	5.1485	-	7470A	PPB	22 Aug 2022 17:28:05	U	SMPL
CCB87 - 1	CCB87	-171	-0.0302	-	7470A	PPB	22 Aug 2022 17:30:22	U	SMPL