



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/23/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:05
In Date: 08/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:50
Out Date: 08/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4285-01	EXXZ3	1	0.92	8.76	9.68	8.38	85.2	
N4285-02	EXXZ4	2	0.93	8.74	9.67	8.32	84.6	
N4285-03	EXXZ5	3	0.92	8.79	9.71	8.39	85.0	
N4285-04	EXXZ6	4	0.92	8.93	9.85	8.47	84.5	
N4285-05	EXXZ7	5	0.91	9.00	9.91	8.2	81.0	
N4285-06	EXXZ8	6	0.91	9.03	9.94	8.29	81.7	
N4285-07	EXXZ9	7	0.91	8.81	9.72	8.39	84.9	
N4285-08	EXY00	8	0.91	9.07	9.98	8.68	85.7	
N4285-09	EXY01	9	0.92	8.96	9.88	8.6	85.7	
N4285-10	EXY02	10	0.93	8.79	9.72	8.43	85.3	
N4285-11	EXY03	11	0.93	8.79	9.72	8.21	82.8	
N4285-12	EXY04	12	0.92	8.77	9.69	8.35	84.7	
N4285-13	EXY05	13	0.93	8.96	9.89	8.34	82.7	
N4285-14	EXY06	14	0.91	8.79	9.7	8.43	85.6	
N4285-15	EXY07	15	0.93	8.79	9.72	8.51	86.2	
N4285-16	EXY08	16	0.91	9.08	9.99	8.62	84.9	
N4285-17	EXY09	17	0.92	8.76	9.68	8.43	85.7	
N4285-18	EXY10	18	0.93	9.02	9.95	8.67	85.8	
N4285-19	EXY11	19	0.91	8.79	9.7	8.5	86.3	
N4285-20	EXY12	20	0.93	9.02	9.95	8.65	85.6	
N4285-21	EXY12D	21	0.93	9.02	9.95	8.65	85.6	
N4285-22	EXY12S	22	0.93	9.02	9.95	8.65	85.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$