



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/23/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:10
In Date: 08/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:55
Out Date: 08/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121636

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4286-01	EXY16	1	0.93	8.97	9.9	8.8	87.7	
N4286-02	EXY17	2	0.92	8.69	9.61	8.45	86.7	
N4286-03	EXY18	3	0.92	8.73	9.65	8.35	85.1	
N4286-04	EXY19	4	0.93	9.06	9.99	9.33	92.7	
N4286-05	EXY20	5	0.92	8.98	9.9	8.58	85.3	
N4286-06	EXY21	6	0.93	9.05	9.98	8.91	88.2	
N4286-07	EXY22	7	0.93	8.82	9.75	8.41	84.8	
N4286-08	EXY23	8	0.93	8.92	9.85	8.37	83.4	
N4286-09	EXY24	9	0.93	8.87	9.8	8.3	83.1	
N4286-10	EXY25	10	0.92	9.06	9.98	8.43	82.9	
N4286-11	EXY26	11	0.92	9.06	9.98	8.45	83.1	
N4286-12	EXY27	12	0.93	8.88	9.81	8.27	82.7	
N4286-13	EXY28	13	0.92	8.95	9.87	8.34	82.9	
N4286-14	EXY29	14	0.93	8.77	9.7	8.22	83.1	
N4286-15	EXY30	15	0.92	9.03	9.95	8.55	84.5	
N4286-16	EXY31	16	0.91	8.84	9.75	8.48	85.6	
N4286-17	EXY32	17	0.92	8.96	9.88	8.68	86.6	
N4286-18	EXY33	18	0.92	8.94	9.86	8.55	85.3	
N4286-19	EXY34	19	0.92	8.84	9.76	8.65	87.4	
N4286-20	EXY38	20	0.92	9.07	9.99	8.45	83.0	
N4286-21	EXY38D	21	0.92	9.07	9.99	8.45	83.0	
N4286-22	EXY38S	22	0.92	9.07	9.99	8.45	83.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$