



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB121637

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/22/2022	15:15
2	CAL2	1	Water	NA	NA	20.7	7.00	08/22/2022	15:16
3	CAL3	1	Water	NA	NA	20.7	10.03	08/22/2022	15:17
4	ICV	1	Water	NA	NA	20.8	6.98	08/22/2022	15:20
5	CCV1	1	Water	NA	NA	20.8	2.01	08/22/2022	15:22
6	N4283-02	1	Solid	20.03	20	24.0	10.25	08/22/2022	15:25
7	N4295-01	1	Solid	20.04	20	23.7	10.16	08/22/2022	15:26
8	N4295-01DUP	1	Solid	20.04	20	23.8	10.17	08/22/2022	15:30
9	CCV2	1	Water	NA	NA	20.8	12.03	08/22/2022	15:32

VB 127637

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N4295 WorkList ID : 162505 Department : Wet-Chemistry Date : 08-22-2022 15:04:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/22/2022	Solid	N4283-02	Corrosivity	Cool 4 deg C	PSEG03	K11	337	08/22/2022	9045D
08/19/2022	Solid	N4295-01	Corrosivity	Cool 4 deg C	XRDS01	K11	XRDS-RECYCLED-SAND-SP-A-COMP	08/19/2022	9045D

Date/Time 08/22/22 15:10
 Raw Sample Received by: JR
 Raw Sample Relinquished by: JR

Date/Time 08/22/22 17:10
 Raw Sample Received by: JR
 Raw Sample Relinquished by: JR