



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121640

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/22/2022	15:20
2	CAL2	1	Water	NA	NA	20.8	7.01	08/22/2022	15:21
3	CAL3	1	Water	NA	NA	20.8	10.03	08/22/2022	15:22
4	ICV	1	Water	NA	NA	20.7	6.98	08/22/2022	15:25
5	CCV1	1	Water	NA	NA	20.7	2.02	08/22/2022	15:30
6	N4294-07	1	Water	NA	NA	24.7	8.60	08/22/2022	15:33
7	N4294-08	1	Water	NA	NA	24.7	9.47	08/22/2022	15:42
8	N4294-08DUP	1	Water	NA	NA	24.8	9.48	08/22/2022	15:43
9	N4294-09	1	Water	NA	NA	24.4	6.37	08/22/2022	15:47
10	CCV2	1	Water	NA	NA	20.7	12.03	08/22/2022	15:50

WORKLIST(Hardcopy Internal Chain)

W3 121640

WorkList Name : CORROSIONITY N4294 V1

WorkList ID : 162507

Department : Wet-Chemistry

Date : 08-22-2022 15:06:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/22/2022	Water	N4294-07	Corrosivity	B	Cool 4 deg C	PSEG03 K11	5643	08/22/2022	9040C
08/22/2022	Water	N4294-08	Corrosivity	B	Cool 4 deg C	PSEG03 K11	5641	08/22/2022	9040C
08/22/2022	Water	N4294-09	Corrosivity	B	Cool 4 deg C	PSEG03 K11	5631	08/22/2022	9040C

Date/Time 08/22/22 15:10
 Raw Sample Received by: JRC
 Raw Sample Relinquished by: JRC

Date/Time 08/22/22 17:20
 Raw Sample Received by: JRC
 Raw Sample Relinquished by: JRC