



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121647

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	83.0	08/23/2022	09:45
2	N4259-04		1	15	100.00	>212.0	>212.0	08/23/2022	10:05
3	N4259-05		1	16	100.00	>212.0	>212.0	08/23/2022	10:20
4	N4259-06		1	15	100.00	>212.0	>212.0	08/23/2022	10:50
5	N4259-07		1	17	100.00	>212.0	>212.0	08/23/2022	11:15
6	N4259-08		1	15	100.00	>212.0	>212.0	08/23/2022	11:40
7	N4264-02		1	17	100.00	>212.0	>212.0	08/23/2022	12:05
8	N4277-02		1	18	100.00	>212.0	>212.0	08/23/2022	12:35
9	N4277-02DUP		1	16	100.00	>212.0	>212.0	08/23/2022	13:05
10	N4277-03		1	15	100.00	>212.0	>212.0	08/23/2022	13:30
11	N4291-01		1	16	100.00	>212.0	>212.0	08/23/2022	14:05
12	N4294-07		1	17	100.00	>212.0	>212.0	08/23/2022	14:35
13	N4294-08		1	17	100.00	>212.0	>212.0	08/23/2022	15:07
14	N4294-09		1	15	100.00	>212.0	>212.0	08/23/2022	15:34

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06