



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/24/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 15:35
In Date: 08/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:21
Out Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121653

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4304-01	EXY88	1	0.92	8.89	9.81	8.3	83.0	
N4304-02	EXY89	2	0.92	8.73	9.65	8.1	82.2	
N4304-03	EXY90	3	0.93	9.02	9.95	8.38	82.6	
N4304-04	EXY91	4	0.92	8.82	9.74	8.66	87.8	
N4304-05	EXY92	5	0.91	9.07	9.98	8.53	84.0	
N4304-06	EXY93	6	0.91	8.87	9.78	8.48	85.3	
N4304-07	EXY94	7	0.92	8.84	9.76	8.61	87.0	
N4304-08	EXY95	8	0.92	8.90	9.82	8.47	84.8	
N4304-09	EXY96	9	0.92	9.07	9.99	8.69	85.7	
N4304-10	EXY97	10	0.91	8.96	9.87	8.66	86.5	
N4304-11	EXY98	11	0.92	8.85	9.77	8.58	86.6	
N4304-12	EXY99	12	0.93	9.05	9.98	8.75	86.4	
N4304-13	EXYA0	13	0.92	9.03	9.95	8.51	84.1	
N4304-14	EXYA1	14	0.92	8.88	9.8	8.33	83.4	
N4304-15	EXYA2	15	0.93	8.69	9.62	8.34	85.3	
N4304-16	EXYA3	16	0.91	8.97	9.88	8.66	86.4	
N4304-17	EXYA4	17	0.93	9.01	9.94	8.68	86.0	
N4304-18	EXYA5	18	0.91	9.05	9.96	8.55	84.4	
N4304-19	EXYA6	19	0.93	8.74	9.67	8.23	83.5	
N4304-20	EXYB0	20	0.92	8.79	9.71	8.25	83.4	
N4304-21	EXYB0D	21	0.92	8.79	9.71	8.25	83.4	
N4304-22	EXYB0S	22	0.92	8.79	9.71	8.25	83.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$