



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/24/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 15:50
In Date: 08/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:27
Out Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121654

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4305-01	EXYA7	1	0.93	9.05	9.98	8.82	87.2	
N4305-02	EXYA8	2	0.93	8.74	9.67	8.15	82.6	
N4305-03	EXYA9	3	0.92	9.07	9.99	8.5	83.6	
N4305-04	EXYB1	4	0.93	8.72	9.65	8.58	87.7	
N4305-05	EXYB2	5	0.94	8.95	9.89	8.66	86.3	
N4305-06	EXYB3	6	0.91	8.80	9.71	8.47	85.9	
N4305-07	EXYB4	7	0.92	8.68	9.6	8.43	86.5	
N4305-08	EXYB5	8	0.91	9.07	9.98	8.83	87.3	
N4305-09	EXYB6	9	0.93	8.85	9.78	8.5	85.5	
N4305-10	EXYB7	10	0.92	8.87	9.79	8.32	83.4	
N4305-11	EXYC9	11	0.91	8.83	9.74	8.05	80.9	
N4305-12	EXYD0	12	0.93	9.05	9.98	8.28	81.2	
N4305-13	EXYD1	13	0.92	8.83	9.75	7.85	78.5	
N4305-14	EXYD2	14	0.94	8.70	9.64	7.87	79.7	
N4305-15	EXYD3	15	0.93	8.85	9.78	7.96	79.4	
N4305-16	EXYD3D	16	0.93	8.85	9.78	7.96	79.4	
N4305-17	EXYD3S	17	0.93	8.85	9.78	7.96	79.4	
N4305-18	EXYD4	18	0.94	8.71	9.65	7.95	80.5	
N4305-19	EXYD5	19	0.93	8.80	9.73	7.86	78.8	
N4305-20	EXYD6	20	0.91	9.04	9.95	8.31	81.9	
N4305-21	EXYD7	21	0.93	8.70	9.63	8.00	81.3	
N4305-22	EXYD8	22	0.92	8.83	9.75	8.89	90.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$