



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/24/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 16:15
In Date: 08/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:33
Out Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121655

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4306-01	EXYC7	1	0.92	8.72	9.64	8.48	86.7	
N4306-02	EXYC8	2	0.93	8.88	9.81	8.19	81.8	
N4306-03	EXYD9	3	0.91	8.88	9.79	8.78	88.6	
N4306-04	EXYE0	4	0.93	9.06	9.99	8.92	88.2	
N4306-05	EXYE1	5	0.92	8.80	9.72	9.01	91.9	
N4306-06	EXYE2	6	0.93	8.69	9.62	8.61	88.4	
N4306-07	EXYE3	7	0.91	8.72	9.63	8.55	87.6	
N4306-08	EXYE4	8	0.93	8.85	9.78	8.75	88.4	
N4306-09	EXYE5	9	0.92	8.87	9.79	8.82	89.1	
N4306-10	EXYE6	10	0.91	8.79	9.7	8.84	90.2	
N4306-11	EXYE7	11	0.93	8.78	9.71	8.86	90.3	
N4306-12	EXYE8	12	0.91	8.84	9.75	8.76	88.8	
N4306-13	EXYE9	13	0.92	8.92	9.84	8.75	87.8	
N4306-14	EXYF0	14	0.92	9.02	9.94	8.69	86.1	
N4306-15	EXYF1	15	0.92	8.90	9.82	8.8	88.5	
N4306-16	EXYF2	16	0.93	8.92	9.85	8.62	86.2	
N4306-17	EXYF3	17	0.91	8.91	9.82	8.63	86.6	
N4306-18	EXYF4	18	0.92	8.80	9.72	8.33	84.2	
N4306-19	EXYF5	19	0.93	8.76	9.69	8.28	83.9	
N4306-20	EXYF6	20	0.93	8.88	9.81	8.43	84.5	
N4306-21	EXYF6D	21	0.93	8.88	9.81	8.43	84.5	
N4306-22	EXYF6S	22	0.93	8.88	9.81	8.43	84.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$