



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121663

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/23/2022	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	08/23/2022	16:26
3	CAL3	1	Water	NA	NA	20.7	10.03	08/23/2022	16:27
4	ICV	1	Water	NA	NA	20.8	6.98	08/23/2022	16:30
5	CCV1	1	Water	NA	NA	20.7	2.01	08/23/2022	16:32
6	N4325-04	1	Water	NA	NA	25.3	8.85	08/23/2022	16:50
7	N4325-04DUP	1	Water	NA	NA	25.4	8.86	08/23/2022	16:51
8	N4325-05	1	Water	NA	NA	23.5	11.30	08/23/2022	17:00
9	N4325-06	1	Water	NA	NA	24.3	7.16	08/23/2022	17:02
10	CCV2	1	Water	NA	NA	20.8	12.03	08/23/2022	17:05

121663

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N4325 w

WorkList ID : 162557

Department : Wet-Chemistry

Date : 08-23-2022 16:44:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/23/2022	Water	N4325-04	Corrosivity	Cool 4 deg C	PSEG03	M11	2632	08/23/2022	9040C
08/23/2022	Water	N4325-05	Corrosivity	Cool 4 deg C	PSEG03	M11	2657	08/23/2022	9040C
08/23/2022	Water	N4325-06	Corrosivity	Cool 4 deg C	PSEG03	M11	2640	08/23/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/23/22 16:10

JD (w)

JD (w)

08/23/22

18:00

JD (w)

JD (w)