



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/25/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 13:10
In Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 06:58
Out Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121665

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4307-01	EXYG1	1	0.91	8.93	9.84	8.33	83.1	
N4307-02	EXYG2	2	0.93	9.05	9.98	8.45	83.1	
N4307-03	EXYG3	3	0.92	8.95	9.87	8.38	83.4	
N4307-04	EXYG4	4	0.93	8.78	9.71	8.17	82.5	
N4307-05	EXYG5	5	0.91	8.95	9.86	8.46	84.4	
N4307-06	EXYG6	6	0.92	8.69	9.61	8.4	86.1	
N4307-07	EXYG7	7	0.91	9.04	9.95	8.55	84.5	
N4307-08	EXYG8	8	0.92	8.93	9.85	8.35	83.2	
N4307-09	EXYG9	9	0.92	9.03	9.95	8.45	83.4	
N4307-10	EXYH0	10	0.92	8.81	9.73	8.09	81.4	
N4307-11	EXYH1	11	0.92	9.05	9.97	8.25	81.0	
N4307-12	EXYH2	12	0.93	8.78	9.71	8.51	86.3	
N4307-13	EXYH3	13	0.92	8.91	9.83	8.48	84.8	
N4307-14	EXYH4	14	0.91	8.79	9.7	8.26	83.6	
N4307-15	EXYH5	15	0.91	9.08	9.99	8.74	86.2	
N4307-16	EXYH6	16	0.92	8.81	9.73	8.47	85.7	
N4307-17	EXYH7	17	0.92	8.72	9.64	8.24	83.9	
N4307-18	EXYH8	18	0.91	9.01	9.92	8.61	85.5	
N4307-19	EXYH9	19	0.91	8.97	9.88	8.65	86.3	
N4307-20	EXYJ3	20	0.93	9.00	9.93	8.37	82.7	
N4307-21	EXYJ3D	21	0.93	9.00	9.93	8.37	82.7	
N4307-22	EXYJ3S	22	0.93	9.00	9.93	8.37	82.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$