



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/25/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 13:30
In Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:10
Out Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121667

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4309-01	EXYM1	1	0.91	9.07	9.98	8.7	85.9	
N4309-02	EXYM2	2	0.92	9.01	9.93	8.53	84.5	
N4309-03	EXYM3	3	0.92	9.07	9.99	8.52	83.8	
N4309-04	EXYM4	4	0.91	8.83	9.74	8.39	84.7	
N4309-05	EXYM5	5	0.92	8.91	9.83	8.73	87.7	
N4309-06	EXYM6	6	0.91	8.91	9.82	8.4	84.1	
N4309-07	EXYM7	7	0.92	9.00	9.92	8.37	82.8	
N4309-08	EXYM8	8	0.91	8.91	9.82	8.2	81.8	
N4309-09	EXYM9	9	0.91	8.77	9.68	8.09	81.9	
N4309-10	EXYN0	10	0.93	8.85	9.78	8.08	80.8	
N4309-11	EXYN1	11	0.93	8.92	9.85	8.71	87.2	
N4309-12	EXYN2	12	0.92	9.06	9.98	8.78	86.8	
N4309-13	EXYN3	13	0.92	9.07	9.99	8.9	88.0	
N4309-14	EXYN4	14	0.92	8.79	9.71	8.8	89.6	
N4309-15	EXYN5	15	0.92	8.93	9.85	8.9	89.4	
N4309-16	EXYN6	16	0.92	8.92	9.84	8.92	89.7	
N4309-17	EXYN7	17	0.91	9.03	9.94	8.89	88.4	
N4309-18	EXYN8	18	0.92	8.73	9.65	8.61	88.1	
N4309-19	EXYN9	19	0.91	8.97	9.88	8.85	88.5	
N4309-20	EXYN9D	20	0.91	8.97	9.88	8.85	88.5	
N4309-21	EXYN9S	21	0.91	8.97	9.88	8.85	88.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$