



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/25/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 13:40
In Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:15
Out Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121668

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4311-01	EXYP0	1	0.92	8.92	9.84	8.39	83.7	
N4311-02	EXYP1	2	0.92	8.95	9.87	8.47	84.4	
N4311-03	EXYP2	3	0.92	8.99	9.91	8.33	82.4	
N4311-04	EXYP3	4	0.91	8.92	9.83	8.69	87.2	
N4311-05	EXYP4	5	0.92	8.83	9.75	8.58	86.7	
N4311-06	EXYP5	6	0.92	8.89	9.81	8.66	87.1	
N4311-07	EXYP6	7	0.92	9.05	9.97	8.81	87.2	
N4311-08	EXYP7	8	0.92	8.85	9.77	8.47	85.3	
N4311-09	EXYP8	9	0.92	8.93	9.85	8.62	86.2	
N4311-10	EXYP9	10	0.93	8.83	9.76	8.81	89.2	
N4311-11	EXYQ0	11	0.93	8.92	9.85	8.87	89.0	
N4311-12	EXYQ1	12	0.93	8.89	9.82	8.52	85.4	
N4311-13	EXYQ2	13	0.93	8.78	9.71	8.35	84.5	
N4311-14	EXYQ3	14	0.93	8.85	9.78	7.86	78.3	
N4311-15	EXYQ4	15	0.91	8.90	9.81	7.84	77.9	
N4311-16	EXYQ5	16	0.92	8.96	9.88	8.08	79.9	
N4311-17	EXYQ6	17	0.93	8.81	9.74	8.38	84.6	
N4311-18	EXYQ7	18	0.91	9.04	9.95	8.35	82.3	
N4311-19	EXYQ8	19	0.91	8.80	9.71	8.24	83.3	
N4311-20	EXYR2	20	0.92	8.78	9.7	8.3	84.1	
N4311-21	EXYR2D	21	0.92	8.78	9.7	8.3	84.1	
N4311-22	EXYR2S	22	0.92	8.78	9.7	8.3	84.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$