



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/25/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 13:50
In Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:27
Out Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121669

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4312-01	EXYS1	1	0.93	8.70	9.63	8.23	83.9	
N4312-02	EXYS2	2	0.93	8.83	9.76	8.22	82.6	
N4312-03	EXYS3	3	0.91	8.96	9.87	8.16	80.9	
N4312-04	EXYS4	4	0.93	8.78	9.71	8.28	83.7	
N4312-05	EXYS5	5	0.92	9.00	9.92	8.44	83.6	
N4312-06	EXYS6	6	0.92	9.03	9.95	8.43	83.2	
N4312-07	EXYS7	7	0.93	9.03	9.96	8.57	84.6	
N4312-08	EXYS8	8	0.92	8.89	9.81	8.25	82.5	
N4312-09	EXYS9	9	0.93	8.97	9.9	8.58	85.3	
N4312-10	EXYT0	10	0.92	9.00	9.92	8.6	85.3	
N4312-11	EXYT1	11	0.92	8.98	9.9	8.59	85.4	
N4312-12	EXYT2	12	0.92	8.81	9.73	8.94	91.0	
N4312-13	EXYT3	13	0.92	8.71	9.63	8.8	90.5	
N4312-14	EXYT4	14	0.93	8.92	9.85	9.07	91.3	
N4312-15	EXYT5	15	0.93	8.95	9.88	9.09	91.2	
N4312-16	EXYT5D	16	0.93	8.95	9.88	9.09	91.2	
N4312-17	EXYT5S	17	0.93	8.95	9.88	9.09	91.2	
N4312-18	EXYT6	18	0.93	9.00	9.93	8.73	86.7	
N4312-19	EXYT7	19	0.92	8.95	9.87	8.79	87.9	
N4312-20	EXYT8	20	0.93	8.74	9.67	8.56	87.3	
N4312-21	EXYT9	21	0.92	8.77	9.69	8.17	82.7	
N4312-22	EXYW0	22	0.92	9.05	9.97	8.28	81.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$