

LB121671


7470A

INSTRUMENT ID: CVR

Linear

 μ Abs.:

47180



A= 0.0000e+000

B= 2.1088e-004

C= -4.4607e-002

Rho= 0.9986706

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.035	-0.035	45	0.000	45					
0.2	0.200	0.170	-0.030	1019	0.0 %	1019					
2.5	2.500	2.384	-0.116	11516	0.0 %	11516					
5.0	5.000	5.416	0.416	25894	0.0 %	25894					
7.5	7.500	7.361	-0.139	35116	0.0 %	35116					
10.0	10.000	9.905	-0.095	47180	0.0 %	47180					

o/b
1-1200516
1-1200517

LB121671 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Aug 2022 09:24:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7470A	PPB	24 Aug 2022 09:47:47	S	Std
0.2 - 1	51.2	1019	-	0.2000	7470A	PPB	24 Aug 2022 09:50:04	S	Std
2.5 - 1	52.5	11516	-	2.5000	7470A	PPB	24 Aug 2022 09:54:56	S	Std
5.0 - 1	55	25894	-	5.0000	7470A	PPB	24 Aug 2022 10:06:34	S	Std
7.5 - 1	57.5	35116	-	7.5000	7470A	PPB	24 Aug 2022 10:08:50	S	Std
10.0 - 1	510	47180	-	10.0000	7470A	PPB	24 Aug 2022 10:11:06	S	Std
ICV02 - 1	ICV02	18996	3.9612	-	7470A	PPB	24 Aug 2022 10:16:36	U	SMPL
ICB02 - 1	ICB02	-58	-0.0568	-	7470A	PPB	24 Aug 2022 10:18:51	U	SMPL
CCV95 - 1	CCV95	23175	4.8425	-	7470A	PPB	24 Aug 2022 10:21:08	U	SMPL
CCB95 - 1	CCB95	-59	-0.0570	-	7470A	PPB	24 Aug 2022 10:23:23	U	SMPL
CRA - 1	CRA	1049	0.1766	-	7470A	PPB	24 Aug 2022 10:25:40	U	SMPL
HighStd - 1	HighStd	47547	9.9820	-	7470A	PPB	24 Aug 2022 10:27:56	U	SMPL
ChkStd - 1	ChkStd	34429	7.2157	-	7470A	PPB	24 Aug 2022 10:33:00	U	SMPL
PB147241BL - 1	PBW	-42	-0.0535	-	7470A	PPB	24 Aug 2022 10:35:15	U	SMPL
PB147241BS - 1	LCSW	16503	3.4355	-	7470A	PPB	24 Aug 2022 10:37:33	U	SMPL
N4272-02 - 1	PSC-124042	-13	-0.0473	-	7470A	PPB	24 Aug 2022 10:39:50	U	SMPL
N4272-04 - 1	PSC-124055	78	-0.0282	-	7470A	PPB	24 Aug 2022 10:42:07	U	SMPL
N4272-06 - 1	PSC-124035	118	-0.0197	-	7470A	PPB	24 Aug 2022 10:44:24	U	SMPL
N4272-08 - 1	PSC-164717	76	-0.0286	-	7470A	PPB	24 Aug 2022 10:46:41	U	SMPL
N4283-02 - 1	337	49	-0.0343	-	7470A	PPB	24 Aug 2022 10:48:56	U	SMPL
CCV96 - 1	CCV96	22611	4.7235	-	7470A	PPB	24 Aug 2022 10:51:11	U	SMPL
CCB96 - 1	CCB96	-65	-0.0583	-	7470A	PPB	24 Aug 2022 10:53:27	U	SMPL
N4283-02DUP - 1	337DUP	23	-0.0398	-	7470A	PPB	24 Aug 2022 10:55:42	U	SMPL
N4283-02MS - 1	337MS	18330	3.8208	-	7470A	PPB	24 Aug 2022 10:57:58	U	SMPL
N4283-02MSD - 1	337MSD	17717	3.6915	-	7470A	PPB	24 Aug 2022 11:00:14	U	SMPL
PB147242BL - 1	PBW	-32	-0.0514	-	7470A	PPB	24 Aug 2022 11:02:29	U	SMPL
PB147242BS - 1	LCSW	16765	3.4907	-	7470A	PPB	24 Aug 2022 11:04:45	U	SMPL
N4294-07 - 1	5643	65	-0.0309	-	7470A	PPB	24 Aug 2022 11:07:02	U	SMPL
N4294-08 - 1	5641	206	-0.0012	-	7470A	PPB	24 Aug 2022 11:09:18	U	SMPL
N4294-09 - 1	5631	184	-0.0058	-	7470A	PPB	24 Aug 2022 11:11:36	U	SMPL
N4294-09DUP - 1	5631DUP	120	-0.0193	-	7470A	PPB	24 Aug 2022 11:13:52	U	SMPL
N4294-09MS - 1	5631MS	16224	3.3767	-	7470A	PPB	24 Aug 2022 11:16:08	U	SMPL
CCV97 - 1	CCV97	23640	4.9405	-	7470A	PPB	24 Aug 2022 11:18:24	U	SMPL
CCB97 - 1	CCB97	-47	-0.0545	-	7470A	PPB	24 Aug 2022 11:20:40	U	SMPL
N4294-09MSD - 1	5631MSD	16528	3.4408	-	7470A	PPB	24 Aug 2022 11:22:55	U	SMPL
PB147175TB - 1	PB147175TB	-15	-0.0478	-	7470A	PPB	24 Aug 2022 11:25:11	U	SMPL
PB147187TB - 1	PB147187TB	47	-0.0347	-	7470A	PPB	24 Aug 2022 11:27:27	U	SMPL
N4283-02LX5 - 1		-60	-0.0573	-	7470A	PPB	24 Aug 2022 11:29:43	U	SMPL
N4283-02A - 1		18481	3.8526	-	7470A	PPB	24 Aug 2022 11:31:58	U	SMPL
N4294-09LX5 - 1		-114	-0.0686	-	7470A	PPB	24 Aug 2022 11:34:14	U	SMPL
N4294-09A - 1		17460	3.6373	-	7470A	PPB	24 Aug 2022 11:36:31	U	SMPL
CCV98 - 1	CCV98	24749	5.1744	-	7470A	PPB	24 Aug 2022 11:38:48	U	SMPL
CCB98 - 1	CCB98	-81	-0.0617	-	7470A	PPB	24 Aug 2022 11:41:04	U	SMPL