



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/25/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 14:20
In Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:43
Out Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121677

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4313-01	EXYW1	1	0.92	8.87	9.79	8.24	82.5	
N4313-02	EXYW2	2	0.92	8.75	9.67	8.43	85.8	
N4313-03	EXYW3	3	0.91	8.66	9.57	8.25	84.8	
N4313-04	EXYW4	4	0.93	8.70	9.63	8.27	84.4	
N4313-05	EXYW5	5	0.92	8.65	9.57	8.64	89.2	
N4313-06	EXYW6	6	0.93	8.81	9.74	8.98	91.4	
N4313-07	EXYW7	7	0.93	8.75	9.68	8.39	85.3	
N4313-08	EXYW8	8	0.93	8.92	9.85	8.44	84.2	
N4313-09	EXYW9	9	0.92	8.90	9.82	8.75	88.0	
N4313-10	EXYX0	10	0.92	8.74	9.66	8.24	83.8	
N4313-11	EXYX1	11	0.93	8.81	9.74	8.25	83.1	
N4313-12	EXYX2	12	0.93	8.84	9.77	8.89	90.0	
N4313-13	EXYX3	13	0.93	8.88	9.81	8.67	87.2	
N4313-14	EXYX4	14	0.92	8.95	9.87	8.79	87.9	
N4313-15	EXYX5	15	0.92	8.96	9.88	8.98	90.0	
N4313-16	EXYX6	16	0.93	8.76	9.69	8.59	87.4	
N4313-17	EXYX7	17	0.93	8.78	9.71	8.56	86.9	
N4313-18	EXYX8	18	0.93	9.04	9.97	8.85	87.6	
N4313-19	EXYX8D	19	0.93	9.04	9.97	8.85	87.6	
N4313-20	EXYX8S	20	0.93	9.04	9.97	8.85	87.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB121677

WorkList Name : %1-n4313 WorkList ID : 162588 Department : Wet-Chemistry Date : 08-24-2022 11:09:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/29/2022	Solid	N4313-01	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW1	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-02	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW2	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-03	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW3	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-04	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW4	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-05	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW5	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-06	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW6	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-07	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW7	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-08	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW8	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-09	Percent Solids	Cool 4 deg C	USEP01	A11	EXYW9	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-10	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX0	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-11	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX1	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-12	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX2	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-13	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX3	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-14	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX4	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-15	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX5	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-16	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX6	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-17	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX7	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-18	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX8	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-19	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX8D	08/19/2022	Chemtech -SO
08/29/2022	Solid	N4313-20	Percent Solids	Cool 4 deg C	USEP01	A11	EXYX8S	08/19/2022	Chemtech -SO

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

08/24/22 12:00
102629
JTC

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

08/24/22 16:00
JTC
JTC