



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 8/25/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121678

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4327-01	WASTE	1	0.91	9.01	9.92	8.85	88.1	
N4327-02	VOC	2	0.93	9.03	9.96	8.58	84.7	
N4327-03	1	3	0.93	9.03	9.96	8.57	84.6	
N4327-04	2	4	0.91	8.74	9.65	8.32	84.8	
N4327-05	3	5	0.91	8.97	9.88	8.42	83.7	
N4327-06	4	6	0.88	8.84	9.72	8.2	82.8	
N4327-07	5	7	0.92	9.02	9.94	8.77	87.0	
N4327-08	6	8	0.91	8.99	9.9	8.61	85.7	
N4328-01	WASTE	9	0.91	8.81	9.72	9.06	92.5	
N4328-02	VOC	10	0.93	8.78	9.71	8.08	81.4	
N4328-03	1	11	0.93	8.75	9.68	9.00	92.2	
N4328-04	2	12	0.91	8.71	9.62	8.74	89.9	
N4328-05	3	13	0.95	8.98	9.93	9.28	92.8	
N4328-06	4	14	0.91	9.05	9.96	9.16	91.2	
N4328-07	5	15	0.94	8.69	9.63	9.06	93.4	
N4328-07D	5DUP	16	0.91	8.74	9.65	9.08	93.5	
N4330-01	OG-SOIL-A	17	0.91	8.63	9.54	9.32	97.5	
N4330-02	OG-SOIL-B	18	0.92	9.04	9.96	9.21	91.7	
N4330-03	OG-SOILCOMP	19	0.91	8.72	9.63	8.83	90.8	
N4338-01	YANNUZZI-PLYMOUTH-PROJECT	20	1.00	1.00	2.00	2.00	100.0	stone sample
N4352-01	11985	21	0.95	8.80	9.75	6.74	65.8	
N4352-02	11985	22	0.93	9.03	9.96	8.69	85.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$