

LB121685


SFAM01.1

INSTRUMENT ID: CY2

Linear

 μ Abs.:

42714



A= 0.0000e+000

B= 2.3330e-004 Slope

C= -4.0197e-003 Intercept

Rho= 0.9999407

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.016	0.016	84	0.000	84					=
0.05	0.050					185	0				3
0.20	0.200	0.205	0.005	897	0.0 %	897					-3
2.50	2.500	2.434	-0.066	10451	0.0 %	10451					1
5.00	5.000	5.035	0.035	21597	0.0 %	21597					2.1
7.50	7.500	7.549	0.049	32374	0.0 %	32374					0
10.0	10.000	9.961	-0.039	42714	0.0 %	42714					NB 8124122

LB121685

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	84	-		0 SFAM01.1	PPB	8/24/2022 14:17	Std
	0.2 S01	897	-		0.2 SFAM01.1	PPB	8/24/2022 14:19	Std
	2.5 S02	10451	-		2.5 SFAM01.1	PPB	8/24/2022 14:22	Std
	5 S03	21597	-		5 SFAM01.1	PPB	8/24/2022 14:24	Std
	7.5 S04	32374	-		7.5 SFAM01.1	PPB	8/24/2022 14:26	Std
	10 S05	42714	-		10 SFAM01.1	PPB	8/24/2022 14:29	Std
ICV003	ICV003	17027	3.9684 -		SFAM01.1	PPB	8/24/2022 14:34	SMPL
ICB003	ICB003	12	-0.0012 -		SFAM01.1	PPB	8/24/2022 14:36	SMPL
CCV001	CCV001	21788	5.0792 -		SFAM01.1	PPB	8/24/2022 14:38	SMPL
CCB001	CCB001	22	0.0011 -		SFAM01.1	PPB	8/24/2022 14:41	SMPL
PB147258BL	PBW258	67	0.0116 -		SFAM01.1	PPB	8/24/2022 14:43	SMPL
N3670-01	MDL-WATER-QT2	278	0.0608 -		SFAM01.1	PPB	8/24/2022 14:48	SMPL
CCV002	CCV002	21157	4.932 -		SFAM01.1	PPB	8/24/2022 14:50	SMPL
CCB002	CCB002	10	-0.0017 -		SFAM01.1	PPB	8/24/2022 14:52	SMPL