



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/25/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 16:40
In Date: 08/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 08:15
Out Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121692

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4334-01	EXZ28	1	0.92	8.79	9.71	8.75	89.1	
N4334-02	EXZ29	2	0.92	9.03	9.95	8.8	87.3	
N4334-03	EXZ30	3	0.93	9.00	9.93	8.58	85.0	
N4334-04	EXZ31	4	0.91	9.00	9.91	8.99	89.8	
N4334-05	EXZ32	5	0.93	8.86	9.79	8.51	85.6	
N4334-06	EXZ33	6	0.92	8.81	9.73	8.44	85.4	
N4334-07	EXZ34	7	0.92	8.89	9.81	9.02	91.1	
N4334-08	EXZ35	8	0.91	8.98	9.89	9.15	91.8	
N4334-09	EXZ36	9	0.91	9.04	9.95	9.33	93.1	
N4334-10	EXZ37	10	0.91	9.04	9.95	8.64	85.5	
N4334-11	EXZ38	11	0.92	9.00	9.92	8.94	89.1	
N4334-12	EXZ39	12	0.92	9.02	9.94	8.72	86.5	
N4334-13	EXZ40	13	0.91	8.89	9.8	8.51	85.5	
N4334-14	EXZ41	14	0.92	8.92	9.84	9.05	91.1	
N4334-15	EXZ42	15	0.91	8.79	9.7	8.6	87.5	
N4334-16	EXZ43	16	0.91	8.73	9.64	8.24	84.0	
N4334-17	EXZ44	17	0.92	8.80	9.72	8.01	80.6	
N4334-18	EXZ45	18	0.92	9.07	9.99	8.41	82.6	
N4334-19	EXZ46	19	0.92	9.05	9.97	8.44	83.1	
N4334-20	EXZ47	20	0.93	8.86	9.79	8.26	82.7	
N4334-21	EXZ47D	21	0.93	8.86	9.79	8.26	82.7	
N4334-22	EXZ47S	22	0.93	8.86	9.79	8.26	82.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$