

LB121697


SFAM01.1

INSTRUMENT ID: EV1

Linear

 μ Abs.:

45908



A= 0.0000e+000

B= 2.2001e-004 slope

C= -5.1708e-002 y-intercept

Rho= 0.9999294

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	0.057	0.057	495	0.000	495					
0.05	0.050					185	0				1
0.20	0.200	0.202	0.002	1151	0.0 %	1151					-2
2.50	2.500	2.438	-0.062	11317	0.0 %	11317					0
5.00	5.000	5.001	0.001	22965	0.0 %	22965					-1
7.50	7.500	7.454	-0.046	34117	0.0 %	34117					0
10.0	10.000	10.048	0.048	45908	0.0 %	45908					

LB121697

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	495	-		0 SFAM01.1	PPB	8/24/2022 15:57	Std
	0.2 S01	1151	-		0.2 SFAM01.1	PPB	8/24/2022 15:59	Std
	2.5 S02	11317	-		2.5 SFAM01.1	PPB	8/24/2022 16:02	Std
	5 S03	22965	-		5 SFAM01.1	PPB	8/24/2022 16:04	Std
	7.5 S04	34117	-		7.5 SFAM01.1	PPB	8/24/2022 16:14	Std
	10 S05	45908	-		10 SFAM01.1	PPB	8/24/2022 16:22	Std
ICV004	ICV004	18615	4.0437 -		SFAM01.1	PPB	8/24/2022 16:24	SMPL
ICB004	ICB004	422	0.0411 -		SFAM01.1	PPB	8/24/2022 16:27	SMPL
CCV003	CCV003	22145	4.8203 -		SFAM01.1	PPB	8/24/2022 16:29	SMPL
CCB003	CCB003	364	0.0284 -		SFAM01.1	PPB	8/24/2022 16:31	SMPL
PB147259BL	PBS259	224	-0.0024 -		SFAM01.1	PPB	8/24/2022 16:36	SMPL
N3670-03	MDL-SOIL-QT2	533	0.0656 -		SFAM01.1	PPB	8/24/2022 16:50	SMPL
CCV004	CCV004	22831	4.9712 -		SFAM01.1	PPB	8/24/2022 16:52	SMPL
CCB004	CCB004	260	0.0055 -		SFAM01.1	PPB	8/24/2022 16:55	SMPL