



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/26/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 13:10
In Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:16
Out Date: 08/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121704

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4336-01	EXZ76	1	0.93	8.82	9.75	8.5	85.8	
N4336-02	EXZ77	2	0.93	8.84	9.77	8.79	88.9	
N4336-03	EXZ78	3	0.93	9.04	9.97	9.06	89.9	
N4336-04	EXZ79	4	0.93	8.83	9.76	8.86	89.8	
N4336-05	EXZ80	5	0.92	8.79	9.71	8.96	91.5	
N4336-06	EXZ81	6	0.92	8.96	9.88	9.11	91.4	
N4336-07	EXZ82	7	0.92	9.03	9.95	8.71	86.3	
N4336-08	EXZ83	8	0.91	8.97	9.88	8.7	86.8	
N4336-09	EXZ84	9	0.91	8.86	9.77	8.58	86.6	
N4336-10	EXZ85	10	0.93	8.89	9.82	8.6	86.3	
N4336-11	EXZ86	11	0.92	8.90	9.82	8.74	87.9	
N4336-12	EXZ87	12	0.91	8.74	9.65	8.56	87.5	
N4336-13	EXZ88	13	0.91	8.94	9.85	8.8	88.3	
N4336-14	EXZ89	14	0.92	8.89	9.81	8.82	88.9	
N4336-15	EXZ90	15	0.92	8.75	9.67	8.49	86.5	
N4336-16	EXZ91	16	0.91	8.78	9.69	8.71	88.8	
N4336-17	EXZ92	17	0.92	8.93	9.85	8.77	87.9	
N4336-18	EXZ93	18	0.91	8.84	9.75	8.64	87.4	
N4336-19	EXZ93D	19	0.91	8.84	9.75	8.64	87.4	
N4336-20	EXZ93S	20	0.91	8.84	9.75	8.64	87.4	
N4336-21	EXZ94	21	0.92	8.87	9.79	8.58	86.4	
N4336-22	EXZ95	22	0.92	8.83	9.75	8.68	87.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$