



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/26/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 13:20
In Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:27
Out Date: 08/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121705

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4337-01	EXZ68	1	0.92	8.85	9.77	8.77	88.7	
N4337-02	EXZ69	2	0.92	9.00	9.92	8.83	87.9	
N4337-03	EXZ71	3	0.92	8.82	9.74	8.91	90.6	
N4337-04	EXZ72	4	0.94	8.98	9.92	8.64	85.7	
N4337-05	EXZ73	5	0.92	8.88	9.8	8.75	88.2	
N4337-06	EXZ74	6	0.92	9.04	9.96	8.49	83.7	
N4337-07	EXZ75	7	0.93	8.88	9.81	8.63	86.7	
N4337-08	EXZA4	8	0.93	8.83	9.76	8.53	86.1	
N4337-09	EXZA5	9	0.93	8.82	9.75	8.84	89.7	
N4337-10	EXZA6	10	0.92	9.07	9.99	9.02	89.3	
N4337-11	EXZA7	11	0.94	8.92	9.86	8.75	87.6	
N4337-12	EXZA8	12	0.93	8.85	9.78	8.82	89.2	
N4337-13	EXZA9	13	0.92	8.96	9.88	8.95	89.6	
N4337-14	EXZB0	14	0.92	8.83	9.75	8.79	89.1	
N4337-15	EXZB1	15	0.92	8.88	9.8	8.76	88.3	
N4337-16	EXZB2	16	0.92	8.84	9.76	8.54	86.2	
N4337-17	EXZB3	17	0.93	8.92	9.85	8.77	87.9	
N4337-18	EXZB4	18	0.93	9.05	9.98	9.14	90.7	
N4337-19	EXZB5	19	0.92	9.03	9.95	9.06	90.1	
N4337-20	EXZB6	20	0.91	8.84	9.75	8.83	89.6	
N4337-21	EXZB6D	21	0.91	8.84	9.75	8.83	89.6	
N4337-22	EXZB6S	22	0.91	8.84	9.75	8.83	89.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$