



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121709

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	08/25/2022	08:35
2	N4330-05		1	12	100.00	>212.0	>212.0	08/25/2022	09:06
3	N4330-05DUP		1	14	100.00	>212.0	>212.0	08/25/2022	09:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121709

Worklist Name : FLASHPOINT-08-25-202

Worklist ID : 162632

Department : Wet-Chemistry

Date : 08-25-2022 08:30:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/31/2022	Water	N4330-05	Flash Point	Cool 4 deg C	PSEG03	M11	OG-WATER	08/24/2022	1010B

Date/Time

8/25/2022 8:30

Raw Sample Received by:

SI (wetchem)

Raw Sample Relinquished by:

CP (son)

Date/Time

8/25/2022 10:00

Raw Sample Received by:

CP (son)

Raw Sample Relinquished by:

SI (wetchem)