

LB121710 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Aug 2022 09:17:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	305	-	0.0000	7470A	PPB	25 Aug 2022 09:58:02	S	Std
0.2 - 1	50.2	1224	-	0.2000	7470A	PPB	25 Aug 2022 10:00:19	S	Std
2.5 - 1	52.5	12536	-	2.5000	7470A	PPB	25 Aug 2022 10:02:36	S	Std
5.0 - 1	55	22855	-	5.0000	7470A	PPB	25 Aug 2022 10:04:53	S	Std
7.5 - 1	57.5	34641	-	7.5000	7470A	PPB	25 Aug 2022 10:07:11	S	Std
10.0 - 1	510	47927	-	10.0000	7470A	PPB	25 Aug 2022 10:17:51	S	Std
ICV05 - 1	ICV05	18769	3.9560	-	7470A	PPB	25 Aug 2022 10:23:12	U	SMPL
ICB05 - 1	ICB05	-125	-0.0685	-	7470A	PPB	25 Aug 2022 10:25:26	U	SMPL
CCV05 - 1	CCV05	23650	4.9956	-	7470A	PPB	25 Aug 2022 10:27:41	U	SMPL
CCB05 - 1	CCB05	-155	-0.0749	-	7470A	PPB	25 Aug 2022 10:29:56	U	SMPL
CRA - 1	CRA	1202	0.2142	-	7470A	PPB	25 Aug 2022 10:32:12	U	SMPL
HighStd - 1	HighStd	47926	10.1664	-	7470A	PPB	25 Aug 2022 10:34:27	U	SMPL
ChkStd - 1	ChkStd	30089	6.3671	-	7470A	PPB	25 Aug 2022 10:36:43	U	SMPL
PB147256BL - 1	PBW	50	-0.0312	-	7470A	PPB	25 Aug 2022 10:39:00	U	SMPL
PB147256BS - 1	LCSW	17133	3.6075	-	7470A	PPB	25 Aug 2022 10:41:17	U	SMPL
N4303-02 - 1	DRUM-1-9	169	-0.0058	-	7470A	PPB	25 Aug 2022 10:43:33	U	SMPL
N4303-02DUP - 1	DRUM-1-9DUP	185	-0.0024	-	7470A	PPB	25 Aug 2022 10:45:49	U	SMPL
N4303-02MS - 1	DRUM-1-9MS	16620	3.4982	-	7470A	PPB	25 Aug 2022 10:48:06	U	SMPL
N4303-02MSD - 1	DRUM-1-9MSD	16151	3.3983	-	7470A	PPB	25 Aug 2022 10:50:23	U	SMPL
N4316-01 - 1	2670	34	-0.0346	-	7470A	PPB	25 Aug 2022 10:52:39	U	SMPL
CCV06 - 1	CCV06	23619	4.9890	-	7470A	PPB	25 Aug 2022 10:54:54	U	SMPL
CCB06 - 1	CCB06	-187	-0.0817	-	7470A	PPB	25 Aug 2022 10:57:10	U	SMPL
N4325-01 - 1	2630	623	0.0909	-	7470A	PPB	25 Aug 2022 10:59:26	U	SMPL
N4325-02 - 1	2591	443	0.0525	-	7470A	PPB	25 Aug 2022 11:01:41	U	SMPL
N4325-03 - 1	2558	248	0.0110	-	7470A	PPB	25 Aug 2022 11:03:57	U	SMPL
PB147257BL - 1	PBW	164	-0.0069	-	7470A	PPB	25 Aug 2022 11:06:12	U	SMPL
PB147257BS - 1	LCSW	15432	3.2452	-	7470A	PPB	25 Aug 2022 11:08:28	U	SMPL
N4325-04 - 1	2632	301	0.0223	-	7470A	PPB	25 Aug 2022 11:10:44	U	SMPL
N4325-05 - 1	2657	244	0.0101	-	7470A	PPB	25 Aug 2022 11:13:01	U	SMPL
N4325-06 - 1	2640	457	0.0555	-	7470A	PPB	25 Aug 2022 11:15:18	U	SMPL
N4325-06DUP - 1	2640DUP	457	0.0555	-	7470A	PPB	25 Aug 2022 11:17:34	U	SMPL
N4325-06MS - 1	2640MS	15321	3.2215	-	7470A	PPB	25 Aug 2022 11:19:50	U	SMPL
CCV07 - 1	CCV07	23257	4.9119	-	7470A	PPB	25 Aug 2022 11:22:06	U	SMPL
CCB07 - 1	CCB07	-165	-0.0770	-	7470A	PPB	25 Aug 2022 11:24:22	U	SMPL
N4325-06MSD - 1	2640MSD	16164	3.4011	-	7470A	PPB	25 Aug 2022 11:29:13	U	SMPL
PB147215TB - 1	PB147215TB	48	-0.0316	-	7470A	PPB	25 Aug 2022 11:31:28	U	SMPL
PB147230TB - 1	PB147230TB	218	0.0046	-	7470A	PPB	25 Aug 2022 11:33:43	U	SMPL
N4303-02LX5 - 1		-47	-0.0519	-	7470A	PPB	25 Aug 2022 11:35:59	U	SMPL
N4303-02A - 1		20987	4.4284	-	7470A	PPB	25 Aug 2022 11:38:15	U	SMPL
N4325-06LX5 - 1		-47	-0.0519	-	7470A	PPB	25 Aug 2022 11:40:31	U	SMPL
N4325-06A - 1		22365	4.7219	-	7470A	PPB	25 Aug 2022 11:42:48	U	SMPL
CCV08 - 1	CCV08	22555	4.7624	-	7470A	PPB	25 Aug 2022 11:45:05	U	SMPL
CCB08 - 1	CCB08	-99	-0.0629	-	7470A	PPB	25 Aug 2022 11:47:23	U	SMPL