

LB121725


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46304



A= 0.0000e+000

B= 2.1804e-004

C= -6.7083e-002

Rho= 0.9997624

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.042	-0.042	117	0.000	117					-
0.05	0.050					185					-
0.20	0.200	0.186	-0.014	1159	0.0 %	1159	0				-7
2.50	2.500	2.545	0.045	11981	0.0 %	11981					2
5.00	5.000	5.121	0.121	23793	0.0 %	23793					2
7.50	7.500	7.361	-0.139	34066	0.0 %	34066					-2
10.0	10.000	10.029	0.029	46304	0.0 %	46304					0

LB121725

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	117	-		0 SFAM01.1	PPB	8/25/2022 14:17	Std
	0.2 S01	1159	-		0.2 SFAM01.1	PPB	8/25/2022 14:19	Std
	2.5 S02	11981	-		2.5 SFAM01.1	PPB	8/25/2022 14:21	Std
	5 S03	23793	-		5 SFAM01.1	PPB	8/25/2022 14:25	Std
	7.5 S04	34066	-		7.5 SFAM01.1	PPB	8/25/2022 14:27	Std
	10 S05	46304	-		10 SFAM01.1	PPB	8/25/2022 14:35	Std
ICV007	ICV007	19250	4.1302	-	SFAM01.1	PPB	8/25/2022 14:40	SMPL
ICB007	ICB007	162	-0.0318	-	SFAM01.1	PPB	8/25/2022 14:45	SMPL
CCV011	CCV011	24285	5.2281	-	SFAM01.1	PPB	8/25/2022 14:47	SMPL
CCB011	CCB011	96	-0.0462	-	SFAM01.1	PPB	8/25/2022 14:49	SMPL
PB147295BL	PBS295	158	-0.0326	-	SFAM01.1	PPB	8/25/2022 14:52	SMPL
N3670-04	MDL-SOIL-QT2	482	0.038	-	SFAM01.1	PPB	8/25/2022 15:03	SMPL
CCV012	CCV012	23536	5.0648	-	SFAM01.1	PPB	8/25/2022 15:05	SMPL
CCB0012	CCB012	138	-0.037	-	SFAM01.1	PPB	8/25/2022 15:08	SMPL