



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 8/26/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 08/25/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 08/26/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB121726

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N4363-01	DBVV6	1	0.93	9.06	9.99	8.18	80.0	
N4363-02	DBVV7	2	0.92	8.87	9.79	7.93	79.0	
N4363-03	DBVV8	3	0.92	8.93	9.85	8.12	80.6	
N4363-04	DBVV9	4	0.92	8.74	9.66	7.94	80.3	
N4363-05	DBVW0	5	0.92	9.06	9.98	8.25	80.9	
N4363-06	DBVW1	6	0.93	8.72	9.65	7.97	80.7	
N4363-07	DBVW2	7	0.92	8.89	9.81	8.46	84.8	
N4363-08	DBVW3	8	0.91	9.04	9.95	8.45	83.4	
N4363-09	DBVW4	9	0.92	8.94	9.86	9.43	95.2	
N4363-10	DBVW5	10	0.93	9.04	9.97	8.43	83.0	
N4363-11	DBVW6	11	0.91	8.98	9.89	8.14	80.5	
N4363-12	DBVW6MS	12	0.91	8.98	9.89	8.14	80.5	
N4363-13	DBVW6MSD	13	0.91	8.98	9.89	8.14	80.5	
N4363-14	DBVW8	14	0.92	8.96	9.88	8.19	81.1	
N4363-15	DBVW9	15	0.92	8.95	9.87	8.09	80.1	
N4363-16	DBVX0	16	0.91	8.77	9.68	7.74	77.9	
N4363-17	DBVX1	17	0.92	9.07	9.99	8.66	85.3	
N4363-18	DBVX2	18	0.91	9.03	9.94	8.47	83.7	
N4363-19	DBVX3	19	0.92	8.69	9.61	8.09	82.5	
N4363-20	DBVX4	20	0.92	8.92	9.84	8.12	80.7	
N4363-21	DBVX5	21	0.92	8.93	9.85	8.00	79.3	
N4363-22	DBVX7	22	0.91	8.92	9.83	7.88	78.1	
N4363-23	VHBLK001	23	1.00	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$