



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 08/25/2022 16:40 TEMP1 OUT: 103 °C 08/26/2022 08:00
TEMP2 IN: 104 °C 08/26/2022 08:30 TEMP2 OUT: 104 °C 08/26/2022 10:00
TEMP3 IN: 540 °C 08/26/2022 10:30 TEMP3 OUT: 540 °C 08/26/2022 12:00
TEMP4 IN: 560 °C 08/26/2022 12:30 TEMP4 OUT: 560 °C 08/26/2022 14:00

Run Number: LB121730

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB121730BL	58.6347	58.6347	100	58.6347	58.6347	58.6347	58.6347	0.0000	0
2	N4371-01	63.8050	63.8050	100	63.8455	63.8455	63.8301	63.8301	0.0154	154
3	N4371-01DUP	89.1406	89.1406	100	89.1825	89.1825	89.1676	89.1676	0.0149	149

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB121730

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs n4371 w

WorkList ID : 162649

Department : Wet-Chemistry

Date : 08-25-2022 15:27:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/30/2022	Water	N4371-01	TVS	Cool 4 deg C	PSEG03	N11	A4127	08/25/2022	SM2540 G

Date/Time

08/25/22 15:30

Raw Sample Received by:

JP (WC)

Raw Sample Relinquished by:

JP (WC)

Date/Time

08/25/22 17:00

Raw Sample Received by:

JP (WC)

Raw Sample Relinquished by:

JP (WC)