

Analytical Summary ReportAnalysis Method: 365.3ANALYST: IwonaParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: JigneshRun Number: LB121731BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP98536
calibration std. phosphate 0.5 ppm	WP98535
calibration std. phosphate 0.3 ppm	WP98534
calibration std. phosphate 0.1 ppm	WP98533
calibration std. phosphate 0.05 ppm	WP98532
calibration std. 0 ppm	WP98531
phosphate CCV std.	WP98537
Combined reagent	WP98643
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphosphate LOQ std, 0.05PPM	WP98541
Phosphate ICV-LCS Std	WP98538
Phosphate LOD-MDL Std 0.025ppm	WP98540

Intercept: -0.0021Slope: 0.6551Regression: 0.999912

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		08/25/2022	14:15
2	CAL2	0.05	1	50	50	0.032	0.052	-4	08/25/2022	14:15
3	CAL3	0.10	1	50	50	0.065	0.102	-2	08/25/2022	14:16
4	CAL4	0.30	1	50	50	0.188	0.29	3.3	08/25/2022	14:16
5	CAL5	0.50	1	50	50	0.325	0.499	0.2	08/25/2022	14:17
6	CAL6	1.00	1	50	50	0.655	1.003	-0.3	08/25/2022	14:17



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB121731

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.498	08/25/2022	14:18
2	ICB		1	50	50	0.001	0.005	08/25/2022	14:18
3	CCV1	0.50	1	50	50	0.327	0.502	08/25/2022	14:35
4	CCB1		1	50	50	0.001	0.005	08/25/2022	14:35
5	PB147296BL		1	1	50	0.002	0.006	08/25/2022	14:36
6	PB147296BS	0.50	1	1	50	0.324	0.498	08/25/2022	14:36
7	N3980-01		1	1	50	0.016	0.028	08/25/2022	14:37
8	N3980-02		1	1	50	0.032	0.052	08/25/2022	14:37
9	N4361-01		1	1.04	50	2.696	4.119	08/25/2022	14:38
10	N4361-01DUP		1	1.02	50	2.670	4.079	08/25/2022	14:38
11	N4361-01MS	0.50	1	1.06	50	2.714	4.146	08/25/2022	14:39
12	N4361-01MSD	0.50	1	1.06	50	2.725	4.163	08/25/2022	14:39
13	N4361-01		20	1.04	50	0.460	0.705	08/25/2022	14:40
14	N4361-01DUP		20	1.02	50	0.461	0.707	08/25/2022	14:40
15	CCV2	0.50	1	50	50	0.322	0.495	08/25/2022	14:41
16	CCB2		1	50	50	0.002	0.006	08/25/2022	14:41
17	N4361-01MS	0.50	20	1.06	50	0.450	0.690	08/25/2022	14:42
18	N4361-01MSD	0.50	20	1.06	50	0.449	0.689	08/25/2022	14:42
19	CCV3	0.50	1	50	50	0.326	0.501	08/25/2022	14:43
20	CCB3		1	50	50	0.003	0.008	08/25/2022	14:43