

LB121732


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37610



A= 0.0000e+000

B= 2.6299e-004

C= 1.1126e-002

Rho= 0.9998428

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% of
0.0	0.000	-0.002	-0.002	-49	0.000	-49					20
0.05	0.050	0.060	0.010	185	0.0 %	185	0				-4
0.2	0.200	0.192	-0.008	689	0.0 %	689					-3
2.5	2.500	2.432	-0.068	9205	0.0 %	9205					1
5.0	5.000	5.049	0.049	19156	0.0 %	19156					2
7.5	7.500	7.617	0.117	28919	0.0 %	28919					1
10.0	10.000	9.902	-0.098	37610	0.0 %	37610					

LB121732

Method: 245.1

Operator: Admin

Date of Analysis: 25 Aug 2022 15:43:10

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-49	-	0.0000	245.1	PPB	25 Aug 2022 15:47:31	S	Std
0.05 - 1	50.05	185	-	0.0500	245.1	PPB	25 Aug 2022 15:49:48	S	Std
0.2 - 1	50.2	689	-	0.2000	245.1	PPB	25 Aug 2022 15:52:05	S	Std
2.5 - 1	52.5	9205	-	2.5000	245.1	PPB	25 Aug 2022 15:54:22	S	Std
5.0 - 1	55	19156	-	5.0000	245.1	PPB	25 Aug 2022 15:58:38	S	Std
7.5 - 1	57.5	28919	-	7.5000	245.1	PPB	25 Aug 2022 16:06:45	S	Std
10.0 - 1	51.8	37810	-	10.0000	245.1	PPB	25 Aug 2022 16:25:54	S	Std
ICV08 - 1	ICV08	14850	3.9165	-245.1		PPB	25 Aug 2022 16:28:39	U	SMPL
ICB08 - 1	ICB08	-79	-0.0097	-245.1		PPB	25 Aug 2022 16:30:54	U	SMPL
CCV13 - 1	CCV13	17665	4.6568	-245.1		PPB	25 Aug 2022 16:33:09	U	SMPL
CCB13 - 1	CCB13	-109	-0.0175	-245.1		PPB	25 Aug 2022 16:35:23	U	SMPL
CRA - 1	CRA	119	0.0424	-245.1		PPB	25 Aug 2022 16:40:12	U	SMPL
HighStd - 1	HighStd	37949	9.9913	-245.1		PPB	25 Aug 2022 16:42:27	U	SMPL
ChkStd - 1	ChkStd	27507	7.2452	-245.1		PPB	25 Aug 2022 16:44:43	U	SMPL
PB147298BL - 1	PBW	-154	-0.0294	-245.1		PPB	25 Aug 2022 16:47:01	U	SMPL
PB147298BS - 1	LCSW	14144	3.7309	-245.1		PPB	25 Aug 2022 16:51:50	U	SMPL
N3980-10 - 1	LOD-MDL-WATER-04-QT3-2022	176	0.0574	-245.1		PPB	25 Aug 2022 16:57:06	U	SMPL
N3980-11 - 1	LOQ-WATER-05-QT3-2022	736	0.2047	-245.1		PPB	25 Aug 2022 16:59:22	U	SMPL
CCV14 - 1	CCV14	20642	5.4398	-245.1		PPB	25 Aug 2022 17:01:39	U	SMPL
CCB14 - 1	CCB14	-152	-0.0288	-245.1		PPB	25 Aug 2022 17:03:56	U	SMPL