



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB121739

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/25/2022	15:35
2	CAL2	1	Water	NA	NA	20.7	7.00	08/25/2022	15:36
3	CAL3	1	Water	NA	NA	20.7	10.03	08/25/2022	15:37
4	ICV	1	Water	NA	NA	20.8	6.98	08/25/2022	15:40
5	CCV1	1	Water	NA	NA	20.7	2.02	08/25/2022	15:42
6	N4364-12	1	Water	NA	NA	27.2	8.35	08/25/2022	15:45
7	N4364-14	1	Water	NA	NA	27.4	8.89	08/25/2022	15:48
8	N4371-01	1	Water	NA	NA	25.6	7.39	08/25/2022	15:50
9	N4371-01DUP	1	Water	NA	NA	25.5	7.40	08/25/2022	15:51
10	N4382-01	1	Water	NA	NA	24.1	5.48	08/25/2022	15:53
11	CCV2	1	Water	NA	NA	20.8	12.03	08/25/2022	16:00

WORKLIST(Hardcopy Internal Chain)

23 1217 39

WorkList Name : ph n4371 w WorkList ID : 162650 Department : Wet-Chemistry Date : 08-25-2022 15:28:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/25/2022	Water	N4364-12	pH	A	Cool 4 deg C	PSEG03	M11 33790	08/25/2022	9040C
08/25/2022	Water	N4364-14	pH	A	Cool 4 deg C	PSEG03	M11 33791	08/25/2022	9040C
08/25/2022	Water	N4371-01	pH	G	Cool 4 deg C	PSEG03	N11 A4127	08/25/2022	9040C
08/23/2022	Water	N4382-01	pH	A	Cool 4 deg C	STER08	M11 POOL-WATER	08/23/2022	9040C

Date/Time 08/25/22 15:30
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM

Date/Time 08/25/22 18:00
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM