



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB121740

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/25/2022	15:35
2	CAL2	1	Water	NA	NA	20.7	7.00	08/25/2022	15:36
3	CAL3	1	Water	NA	NA	20.7	10.03	08/25/2022	15:37
4	ICV	1	Water	NA	NA	20.8	6.98	08/25/2022	15:40
5	CCV1	1	Water	NA	NA	20.7	2.01	08/25/2022	16:05
6	N4364-01	1	Solid	20.02	20	23.1	9.10	08/25/2022	16:08
7	N4364-02	1	Solid	20.03	20	22.4	8.92	08/25/2022	16:10
8	N4364-03	1	Solid	20.04	20	23.1	7.19	08/25/2022	16:11
9	N4364-04	1	Solid	20.02	20	21.7	7.64	08/25/2022	16:15
10	N4364-05	1	Solid	20.03	20	22.2	8.13	08/25/2022	16:20
11	N4364-06	1	Solid	20.05	20	21.9	7.07	08/25/2022	16:22
12	N4364-07	1	Solid	20.03	20	23.2	9.20	08/25/2022	16:25
13	N4364-08	1	Solid	20.02	20	23.1	7.57	08/25/2022	16:30
14	N4374-02	1	Solid	20.02	20	21.6	7.62	08/25/2022	16:32
15	N4374-02DUP	1	Solid	20.03	20	21.7	7.63	08/25/2022	16:33
16	CCV2	1	Water	NA	NA	20.8	12.02	08/25/2022	16:35

WORKLIST(Hardcopy Internal Chain)

VB 121740

WorkList Name : ph s n4371

WorkList ID : 162651

Department : Wet-Chemistry

Date : 08-25-2022 15:29:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/25/2022	Solid	N4364-01	pH	Cool 4 deg C	PSEG03	M11	33793	08/25/2022	9045D
08/25/2022	Solid	N4364-02	pH	Cool 4 deg C	PSEG03	M11	33794	08/25/2022	9045D
08/25/2022	Solid	N4364-03	pH	Cool 4 deg C	PSEG03	M11	33795	08/25/2022	9045D
08/25/2022	Solid	N4364-04	pH	Cool 4 deg C	PSEG03	M11	33796	08/25/2022	9045D
08/25/2022	Solid	N4364-05	pH	Cool 4 deg C	PSEG03	M11	33797	08/25/2022	9045D
08/25/2022	Solid	N4364-06	pH	Cool 4 deg C	PSEG03	M11	33798	08/25/2022	9045D
08/25/2022	Solid	N4364-07	pH	Cool 4 deg C	PSEG03	M11	33799	08/25/2022	9045D
08/25/2022	Solid	N4364-08	pH	Cool 4 deg C	PSEG03	M11	33800	08/25/2022	9045D
08/25/2022	Solid	N4374-02	pH C	Cool 4 deg C	PSEG03	M11	1754	08/25/2022	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/25/22 15:30

JD (CSC)

JD (CSC)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

08/25/22

JD (CSC)

JD (CSC)

18:00

JD (CSC)

JD (CSC)