



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121742

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	08/26/2022	08:35
2	N4364-12		1	15	100.00	>212.0	>212.0	08/26/2022	08:50
3	N4364-14		1	15	100.00	>212.0	>212.0	08/26/2022	10:20
4	N4371-01		1	18	100.00	>212.0	>212.0	08/26/2022	10:45
5	N4371-01DUP		1	18	100.00	>212.0	>212.0	08/26/2022	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FLASHPOINT-08-26-22

WorkList ID : 162654

Department : Wet-Chemistry

Date : 08-26-2022 08:26:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/01/2022	Water	N4364-12	Flash Point	Cool 4 deg C	PSEG03	M11	33790	08/25/2022	1010B
09/01/2022	Water	N4364-14	Flash Point	Cool 4 deg C	PSEG03	M11	33791	08/25/2022	1010B
08/30/2022	Water	N4371-01	Flash Point	Cool 4 deg C	PSEG03	N11	A4127	08/25/2022	1010B

Date/Time 8/26/2022 8:30
 Raw Sample Received by: SJ (wetchem)
 Raw Sample Relinquished by: CP(SM)

Date/Time 8/26/2022 11:30
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: SJ (wetchem)