



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB121744

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP98674
Turbidity Calibration std, 0NTU	WP98676
Turbidity Calibration std, 20NTU	WP98669
Turbidity Calibration std, 40NTU	WP98668
Turbidity Calibration std, 1NTU	WP98672
Turbidity Calibration std, 5NTU	WP98671
Turbidity LOQ std, 1.0NTU	WP98675
Turbidity Calibration - CCV std, 10 NTU	WP98670
10 NTU Standard 500 ml	W2752

Intercept: 0.0794

Slope: 0.9317

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.07		08/26/2022	10:30
2	CAL2	1	1	1.140	1.14	-13.8	08/26/2022	10:33
3	CAL3	5	1	4.460	4.70	6	08/26/2022	10:36
4	CAL4	10	1	9.540	10.15	-1.5	08/26/2022	10:40
5	CAL5	20	1	18.600	19.88	0.6	08/26/2022	10:43
6	CAL6	40	1	37.400	40.06	-0.1	08/26/2022	10:46



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB121744

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.330	9.929	08/26/2022	10:50
2	ICB		1	0.134	0.059	08/26/2022	10:53
3	CCV1	10	1	9.360	9.961	08/26/2022	10:56
4	CCB1		1	0.148	0.074	08/26/2022	11:00
5	LB121744BL		1	0.152	0.078	08/26/2022	11:03
6	N3980-07		1	0.525	0.478	08/26/2022	11:06
7	N3980-08		1	1.130	1.128	08/26/2022	11:10
8	CCV2	10	1	9.240	9.832	08/26/2022	11:13
9	CCB2		1	0.146	0.071	08/26/2022	11:16