

LB121745


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

54688



A= 0.0000e+000

B= 1.8120e-004

C= -2.6402e-002

Rho= 0.9997438

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/20
0.0	0.000	0.004	0.004	170	0.000	170					-
0.2	0.200	0.181	-0.019	1147	0.0 %	1147					-10
2.5	2.500	2.475	-0.025	13804	0.0 %	13804					-1
5.0	5.000	4.991	-0.009	27690	0.0 %	27690					0
7.5	7.500	7.665	0.165	42445	0.0 %	42445					2
10.0	10.000	9.883	-0.117	54688	0.0 %	54688					-1

LB121745 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Aug 2022 10:09:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	170	-	0.0000	7471B	PPB	26 Aug 2022 10:16:40	S	Std
0.2 - 1	50.2	1147	-	0.2000	7471B	PPB	26 Aug 2022 10:18:56	S	Std
2.5 - 1	52.5	13804	-	2.5000	7471B	PPB	26 Aug 2022 10:21:13	S	Std
5.0 - 1	55	27690	-	5.0000	7471B	PPB	26 Aug 2022 10:23:30	S	Std
7.5 - 1	57.5	42445	-	7.5000	7471B	PPB	26 Aug 2022 10:25:47	S	Std
10.0 - 1	511	54688	-	10.0000	7471B	PPB	26 Aug 2022 10:28:04	S	Std
ICV09 - 1	ICV09	23089	4.1574	-7471B	PPB	26 Aug 2022 10:32:16	U	SMPL	
ICB09 - 1	ICB09	-146	-0.0529	-7471B	PPB	26 Aug 2022 10:34:31	U	SMPL	
CCV15 - 1	CCV15	28174	5.0788	-7471B	PPB	26 Aug 2022 10:36:48	U	SMPL	
CCB15 - 1	CCB15	-154	-0.0543	-7471B	PPB	26 Aug 2022 10:39:03	U	SMPL	
CRA - 1	CRA	960	0.1476	-7471B	PPB	26 Aug 2022 10:41:20	U	SMPL	
HighStd - 1	HighStd	55839	10.0918	-7471B	PPB	26 Aug 2022 10:43:35	U	SMPL	
ChkStd - 1	ChkStd	40728	7.3537	-7471B	PPB	26 Aug 2022 10:45:50	U	SMPL	
PB147288BL - 1	PBS	-246	-0.0710	-7471B	PPB	26 Aug 2022 10:48:09	U	SMPL	
PB147288BS - 1	LCSS	20746	3.7329	-7471B	PPB	26 Aug 2022 10:50:26	U	SMPL	
N4303-01 - 1	DRUM-1-9	1001	0.1550	-7471B	PPB	26 Aug 2022 10:57:22	U	SMPL	
N4303-01DUP - 1	DRUM-1-9DUP	1137	0.1796	-7471B	PPB	26 Aug 2022 10:59:38	U	SMPL	
N4303-01MS - 1	DRUM-1-9MS	20680	3.7209	-7471B	PPB	26 Aug 2022 11:01:55	U	SMPL	
N4303-01MSD - 1	DRUM-1-9MSD	20348	3.6607	-7471B	PPB	26 Aug 2022 11:04:12	U	SMPL	
N4310-01 - 1	EO-02-082322	6578	1.1656	-7471B	PPB	26 Aug 2022 11:06:30	U	SMPL	
CCV16 - 1	CCV16	27634	4.9810	-7471B	PPB	26 Aug 2022 11:08:47	U	SMPL	
CCB16 - 1	CCB16	-146	-0.0529	-7471B	PPB	26 Aug 2022 11:11:02	U	SMPL	
N4323-01 - 1	TR-05-082322	2014	0.3385	-7471B	PPB	26 Aug 2022 11:13:20	U	SMPL	
N4326-01 - 1	SU-03-082322	1591	0.2619	-7471B	PPB	26 Aug 2022 11:15:35	U	SMPL	
N4338-01 - 1	YANNUZZI-PLYMOUTH-PROJECT	414	0.0486	-7471B	PPB	26 Aug 2022 11:17:51	U	SMPL	
N4352-02 - 1	11985	1845	0.3079	-7471B	PPB	26 Aug 2022 11:20:06	U	SMPL	
N4353-03 - 1	1-2-COMP-1	280	0.0243	-7471B	PPB	26 Aug 2022 11:22:22	U	SMPL	
N4353-04 - 1	3-4-COMP-2	1072	0.1678	-7471B	PPB	26 Aug 2022 11:24:39	U	SMPL	
N4375-01 - 1	WC-1	1752	0.2911	-7471B	PPB	26 Aug 2022 11:26:55	U	SMPL	
N4375-04 - 1	WC-2	2014	0.3385	-7471B	PPB	26 Aug 2022 11:29:12	U	SMPL	
N4375-07 - 1	WC-3	2956	0.5092	-7471B	PPB	26 Aug 2022 11:31:29	U	SMPL	
N4375-10 - 1	WC-4	6715	1.1904	-7471B	PPB	26 Aug 2022 11:33:45	U	SMPL	
CCV17 - 1	CCV17	26957	4.8583	-7471B	PPB	26 Aug 2022 11:36:00	U	SMPL	
CCB17 - 1	CCB17	-131	-0.0501	-7471B	PPB	26 Aug 2022 11:38:16	U	SMPL	
N4378-01 - 1	CL-01-082522	2049	0.3449	-7471B	PPB	26 Aug 2022 11:40:33	U	SMPL	
N4303-01LX5 - 1		79	-0.0121	-7471B	PPB	26 Aug 2022 11:42:48	U	SMPL	
N4303-01A - 1		22216	3.9992	-7471B	PPB	26 Aug 2022 11:45:04	U	SMPL	
CCV18 - 1	CCV18	27631	4.9804	-7471B	PPB	26 Aug 2022 11:47:20	U	SMPL	
CCB18 - 1	CCB18	-150	-0.0536	-7471B	PPB	26 Aug 2022 11:49:38	U	SMPL	