

LB121752


245.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37023



A= 0.0000e+000

B= 2.6991e-004

C= 1.1003e-002

Rho= 0.9999783

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.0	0.000	-0.018	-0.018	-107	0.000	-107					
0.05	0.050	0.057	0.007	169	0.0 %	169	0				14
0.2	0.200	0.197	-0.003	688	0.0 %	688					-2
2.5	2.500	2.500	0.000	9223	0.0 %	9223					0
5.0	5.000	5.048	0.048	18663	0.0 %	18663					1
7.5	7.500	7.462	-0.038	27605	0.0 %	27605					1
10.0	10.000	10.004	0.004	37023	0.0 %	37023					0

LB121752 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 26 Aug 2022 12:47:00

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-107	-	0.0000	245.1	PPB	26 Aug 2022 13:24:32	S	Std
0.05 - 1	50.2	169	-	0.0500	245.1	PPB	26 Aug 2022 13:26:49	S	Std
0.2 - 1	52.5	688	-	0.2000	245.1	PPB	26 Aug 2022 13:31:50	S	Std
2.5 - 1	55	9223	-	2.5000	245.1	PPB	26 Aug 2022 13:34:07	S	Std
5.0 - 1	57.5	18663	-	5.0000	245.1	PPB	26 Aug 2022 13:36:24	S	Std
7.5 - 1	57.5	27605	-	7.5000	245.1	PPB	26 Aug 2022 13:38:39	S	Std
10.0 - 1	51.1	37023	-	10.0000	245.1	PPB	26 Aug 2022 14:00:31	S	Std
ICV10 - 1	ICV10	14607	3.9536	-245.1		PPB	26 Aug 2022 14:03:15	U	SMPL
ICB10 - 1	ICB10	-149	-0.0292	-245.1		PPB	26 Aug 2022 14:05:30	U	SMPL
CCV19 - 1	CCV19	19135	5.1758	-245.1		PPB	26 Aug 2022 14:07:45	U	SMPL
CCB19 - 1	CCB19	-139	-0.0265	-245.1		PPB	26 Aug 2022 14:10:00	U	SMPL
CRA - 1	CRA	132	0.0486	-245.1		PPB	26 Aug 2022 14:14:51	U	SMPL
HighStd - 1	HighStd	38309	10.3510	-245.1		PPB	26 Aug 2022 14:17:06	U	SMPL
ChkStd - 1	ChkStd	26380	7.1313	-245.1		PPB	26 Aug 2022 14:22:03	U	SMPL
PB147307BL - 1	PBW	-117	-0.0206	-245.1		PPB	26 Aug 2022 14:26:59	U	SMPL
PB147307BS - 1	LCSW	15369	4.1593	-245.1		PPB	26 Aug 2022 14:29:14	U	SMPL
N3980-10 - 1	LOD-MDL-WATER-04-QT3-2022	131	0.0464	-245.1		PPB	26 Aug 2022 14:34:07	U	SMPL
N3980-10RE - 1	LOD-MDL-WATER-04-QT3-2022	531	0.1543	-245.1		PPB	26 Aug 2022 14:36:23	U	SMPL
N3980-12 - 1	MDL-WATER-06-QT3-2022	155	0.0528	-245.1		PPB	26 Aug 2022 14:38:40	U	SMPL
N4264-02 - 1	U-3	-93	-0.0141	-245.1		PPB	26 Aug 2022 14:40:57	U	SMPL
N4277-02 - 1	B-1	1498	0.4153	-245.1		PPB	26 Aug 2022 14:43:12	U	SMPL
CCV20 - 1	CCV20	19603	5.3021	-245.1		PPB	26 Aug 2022 14:45:28	U	SMPL
CCB20 - 1	CCB20	-126	-0.0230	-245.1		PPB	26 Aug 2022 14:47:43	U	SMPL
N4277-02DUP - 1	B-1DUP	1647	0.4555	-245.1		PPB	26 Aug 2022 14:49:58	U	SMPL
N4277-02MS - 1	B-1MS	18078	4.3506	-245.1		PPB	26 Aug 2022 14:52:14	U	SMPL
N4277-02MSD - 1	B-1MSD	17218	4.6583	-245.1		PPB	26 Aug 2022 14:54:29	U	SMPL
N4277-03 - 1	U-2	-88	-0.0127	-245.1		PPB	26 Aug 2022 14:56:45	U	SMPL
N4291-01 - 1	Y-7	-90	-0.0133	-245.1		PPB	26 Aug 2022 14:59:01	U	SMPL
N4277-02LX5 - 1		201	0.0653	-245.1		PPB	26 Aug 2022 15:01:17	U	SMPL
N4277-02A - 1		16809	4.5479	-245.1		PPB	26 Aug 2022 15:03:34	U	SMPL
CCV21 - 1	CCV21	20076	5.4297	-245.1		PPB	26 Aug 2022 15:05:51	U	SMPL
CCB21 - 1	CCB21	-127	-0.0233	-245.1		PPB	26 Aug 2022 15:08:08	U	SMPL