



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB121754

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.1
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/26/2022	09:25
2	CAL2	1	Water	NA	NA	20.8	7.00	08/26/2022	09:26
3	CAL3	1	Water	NA	NA	20.8	10.03	08/26/2022	09:27
4	ICV	1	Water	NA	NA	20.7	6.99	08/26/2022	09:30
5	CCV1	1	Water	NA	NA	20.7	2.01	08/26/2022	09:32
6	N4383-01	1	Water	NA	NA	21.4	7.97	08/26/2022	09:35
7	N4383-02	1	Water	NA	NA	23.1	8.13	08/26/2022	09:38
8	N4385-01	1	Water	NA	NA	23.0	7.88	08/26/2022	09:39
9	N4385-02	1	Water	NA	NA	22.9	7.87	08/26/2022	09:41
10	N4387-01	1	Water	NA	NA	22.6	7.91	08/26/2022	09:43
11	N4387-02	1	Water	NA	NA	22.7	7.92	08/26/2022	09:45
12	N4389-01	1	Water	NA	NA	23.2	7.89	08/26/2022	09:47
13	N4389-02	1	Water	NA	NA	23.5	7.94	08/26/2022	09:49
14	N4391-01	1	Water	NA	NA	23.2	7.53	08/26/2022	09:50
15	N4391-02	1	Water	NA	NA	22.7	7.68	08/26/2022	09:52
16	CCV2	1	Water	NA	NA	20.8	12.02	08/26/2022	09:54
17	N4391-02DUP	1	Water	NA	NA	22.8	7.69	08/26/2022	09:56
18	CCV3	1	Water	NA	NA	20.7	2.01	08/26/2022	09:58

VB 121753-70
VB 121753

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n4383

WorkList ID : 162694

Department : Wet-Chemistry

Date : 08-26-2022 09:14:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/24/2022	Water	N4383-01	pH	Cool 4 deg C	ENV/70	M11	252672-1	08/24/2022	9040C
08/24/2022	Water	N4383-02	pH	Cool 4 deg C	ENV/70	M11	252672-2	08/24/2022	9040C
08/24/2022	Water	N4385-01	pH	Cool 4 deg C	ENV/70	M11	265751-1	08/24/2022	9040C
08/24/2022	Water	N4385-02	pH	Cool 4 deg C	ENV/70	M11	265751-2	08/24/2022	9040C
08/24/2022	Water	N4387-01	pH	Cool 4 deg C	ENV/70	M11	251563-1	08/24/2022	9040C
08/24/2022	Water	N4387-02	pH	Cool 4 deg C	ENV/70	M11	251563-2	08/24/2022	9040C
08/24/2022	Water	N4389-01	pH	Cool 4 deg C	ENV/70	M11	265393-3	08/24/2022	9040C
08/24/2022	Water	N4389-02	pH	Cool 4 deg C	ENV/70	M11	265393-4	08/24/2022	9040C
08/24/2022	Water	N4391-01	pH	Cool 4 deg C	ENV/70	M11	256317-3	08/24/2022	9040C
08/24/2022	Water	N4391-02	pH	Cool 4 deg C	ENV/70	M11	256317-4	08/24/2022	9040C

Date/Time 08/26/22 09:20

Raw Sample Received by: 70609

Raw Sample Relinquished by: [Signature]

Date/Time 08/26/22 12:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: 10609