



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: jignesh

Run Number: LB121770

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP98703
Chlorine Calibration std, 0.1ppm	WP98698
Chlorine Calibration std, 0.2ppm	WP98699
Chlorine Calibration std, 0.8ppm	WP98701
Chlorine Calibration std, 0.0ppm	WP98697
Chlorine Calibration std, 1.6ppm	WP98702
Residual Chlorine Calibration and CCV std, 0	WP98700
Total Chlorine Powder Pillows	W2759

Intercept: 0.0114

Slope: 1.0134

Regression: 0.999903

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		08/26/2022	10:30
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	3	08/26/2022	10:33
3	CAL3	0.2	1	0.000	0.220	0.220	0.21	-3	08/26/2022	10:36
4	CAL4	0.4	1	0.000	0.430	0.430	0.41	-3.2	08/26/2022	10:39
5	CAL5	0.8	1	0.000	0.820	0.820	0.80	0.3	08/26/2022	10:42
6	CAL6	1.6	1	0.000	1.630	1.630	1.60	0.2	08/26/2022	10:45



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: jignesh

Run Number: LB121770

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4200	0.4200	0.4030	08/26/2022	10:48
2	ICB				1	0.0000	0.0000	0.0000	-0.0110	08/26/2022	10:51
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.3930	08/26/2022	10:54
4	CCB1				1	0.0000	0.0100	0.0100	-0.0010	08/26/2022	11:59
5	LB121770BL	50	50		1	0.0000	0.0100	0.0100	-0.0010	08/26/2022	11:02
6	LB121770BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.4030	08/26/2022	11:05
7	N4393-01	50	50		1	0.0000	0.0270	0.0270	0.0150	08/26/2022	11:08
8	N4393-01DUP	50	50		1	0.0000	0.0280	0.0280	0.0160	08/26/2022	11:11
9	CCV2			0.4	1	0.0000	0.4200	0.4200	0.4030	08/26/2022	11:14
10	CCB2				1	0.0000	0.0000	0.0000	-0.0110	08/26/2022	11:17