

LB121771


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4354E



A= 0.0000e+000

B= 2.2429e-004

C= 1.1387e-001

Rho= 0.9990803

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.099	0.099	-68	0.000	-68				
0.2	0.200	0.244	0.044	580	0.0 %	580				
2.5	2.500	2.207	-0.293	9332	0.0 %	9332				
5.0	5.000	5.103	0.103	22243	0.0 %	22243				
7.5	7.500	7.667	0.167	33676	0.0 %	33676				
10.0	10.000	9.881	-0.119	43546	0.0 %	43546				

0/01
 22
 -12
 2
 2
 -1
 AB
 8129123

LB121771 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Aug 2022 11:35:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-68	-	0.0000	7470A	PPB	29 Aug 2022 13:27:29	S	Std
0.2 - 1	50.2	580	-	0.2000	7470A	PPB	29 Aug 2022 13:29:46	S	Std
2.5 - 1	52.5	9332	-	2.5000	7470A	PPB	29 Aug 2022 13:32:03	S	Std
5.0 - 1	55	22243	-	5.0000	7470A	PPB	29 Aug 2022 13:34:20	S	Std
7.5 - 1	57.5	33676	-	7.5000	7470A	PPB	29 Aug 2022 13:36:37	S	Std
10.0 - 1	510	43546	-	10.0000	7470A	PPB	29 Aug 2022 13:41:29	S	Std
ICV11 - 1	ICV11	17093	3.9476	-	7470A	PPB	29 Aug 2022 13:48:47	U	SMPL
ICB11 - 1	ICB11	-108	0.0897	-	7470A	PPB	29 Aug 2022 13:51:02	U	SMPL
CCV22 - 1	CCV22	22184	5.0895	-	7470A	PPB	29 Aug 2022 13:53:17	U	SMPL
CCB22 - 1	CCB22	-186	0.0722	-	7470A	PPB	29 Aug 2022 13:55:32	U	SMPL
CRA - 1	CRA	611	0.2509	-	7470A	PPB	29 Aug 2022 14:00:20	U	SMPL
HighStd - 1	HighStd	42759	9.7042	-	7470A	PPB	29 Aug 2022 14:02:34	U	SMPL
ChkStd - 1	ChkStd	31859	7.2595	-	7470A	PPB	29 Aug 2022 14:04:49	U	SMPL
PB147338BL - 1	PBW	-176	0.0744	-	7470A	PPB	29 Aug 2022 14:07:06	U	SMPL
PB147338BS - 1	LCSW	17848	4.1170	-	7470A	PPB	29 Aug 2022 14:09:24	U	SMPL
N4330-05 - 1	OG-WATER	75	0.1307	-	7470A	PPB	29 Aug 2022 14:16:10	U	SMPL
N4330-05DUP - 1	OG-WATERDUP	37	0.1222	-	7470A	PPB	29 Aug 2022 14:18:26	U	SMPL
N4330-05MS - 1	OG-WATERMS	15858	3.6706	-	7470A	PPB	29 Aug 2022 14:20:43	U	SMPL
N4330-05MSD - 1	OG-WATERMSD	15918	3.6841	-	7470A	PPB	29 Aug 2022 14:23:01	U	SMPL
N4371-01 - 1	A4127	-29	0.1074	-	7470A	PPB	29 Aug 2022 14:25:16	U	SMPL
CCV23 - 1	CCV23	21984	5.0446	-	7470A	PPB	29 Aug 2022 14:27:32	U	SMPL
CCB23 - 1	CCB23	-196	0.0699	-	7470A	PPB	29 Aug 2022 14:29:48	U	SMPL
N4379-21 - 1	EQUIP-BLANK	-151	0.0800	-	7470A	PPB	29 Aug 2022 14:32:03	U	SMPL
N4330-05LX5 - 1		-198	0.0695	-	7470A	PPB	29 Aug 2022 14:41:06	U	SMPL
N4330-05A - 1		18782	4.3265	-	7470A	PPB	29 Aug 2022 14:43:21	U	SMPL
CCV24 - 1	CCV24	21840	5.0123	-	7470A	PPB	29 Aug 2022 14:45:37	U	SMPL
CCB24 - 1	CCB24	-116	0.0879	-	7470A	PPB	29 Aug 2022 14:47:52	U	SMPL