



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB121783

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	08/29/2022	13:05
2	CAL2	1	Water	NA	NA	20.7	7.00	08/29/2022	13:06
3	CAL3	1	Water	NA	NA	20.7	10.03	08/29/2022	13:09
4	ICV	1	Water	NA	NA	20.8	7.02	08/29/2022	13:10
5	CCV1	1	Water	NA	NA	20.8	2.01	08/29/2022	13:11
6	N4421-02	1	Solid	20.03	20	23.8	10.30	08/29/2022	13:15
7	N4421-02DUP	1	Solid	20.04	20	23.7	10.31	08/29/2022	13:16
8	N4421-07	1	Solid	20.02	20	23.7	10.86	08/29/2022	13:19
9	N4421-12	1	Solid	20.03	20	23.7	9.31	08/29/2022	13:22
10	N4423-02	1	Solid	20.04	20	23.7	9.14	08/29/2022	13:25
11	N4423-05	1	Solid	20.02	20	23.8	8.93	08/29/2022	13:30
12	CCV2	1	Water	NA	NA	20.8	12.03	08/29/2022	13:35

121783

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N4421 WorkList ID : 162741 Department : Wet-Chemistry Date : 08-29-2022 12:22:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/26/2022	Solid	N4421-02	Corrosivity	Cool 4 deg C	ROUX02	M11	WC-PALM-SP-01	08/26/2022	9045D
08/26/2022	Solid	N4421-07	Corrosivity	Cool 4 deg C	ROUX02	M11	WC-PALM-SP-02	08/26/2022	9045D
08/26/2022	Solid	N4421-12	Corrosivity	Cool 4 deg C	ROUX02	M11	WC-PALM-SP-03	08/26/2022	9045D
08/27/2022	Solid	N4423-02	Corrosivity	Cool 4 deg C	PSEG03	M11	WC-1	08/27/2022	9045D
08/27/2022	Solid	N4423-05	Corrosivity	Cool 4 deg C	PSEG03	M11	WC-2	08/27/2022	9045D

Date/Time 08/29/22 13:00
 Raw Sample Received by: JRC
 Raw Sample Relinquished by: JRC

Date/Time 08/29/22 14:30
 Raw Sample Received by: JRC
 Raw Sample Relinquished by: JRC