

LB121786


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56820



A= 0.0000e+000

B= 1.7465e-004

C= 1.7723e-002

Rho= 0.9999108

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.022	0.022	25	0.000	25					
0.2	0.200	0.169	-0.031	869	0.0 %	869					-16
2.5	2.500	2.498	-0.002	14203	0.0 %	14203					0
5.0	5.000	4.974	-0.026	28375	0.0 %	28375					-1
7.5	7.500	7.595	0.095	43385	0.0 %	43385					-1
10.0	10.000	9.942	-0.058	56820	0.0 %	56820					-1

LB121786 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Aug 2022 09:34:12

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	7471B	PPB	30 Aug 2022 10:11:34	S	Std
0.2 - 1	50.2	869	-	0.2000	7471B	PPB	30 Aug 2022 10:13:51	S	Std
2.5 - 1	52.5	14203	-	2.5000	7471B	PPB	30 Aug 2022 10:16:08	S	Std
5.0 - 1	55	28375	-	5.0000	7471B	PPB	30 Aug 2022 10:18:25	S	Std
7.5 - 1	57.5	43385	-	7.5000	7471B	PPB	30 Aug 2022 10:20:42	S	Std
10.0 - 1	511	56820	-	10.0000	7471B	PPB	30 Aug 2022 10:22:59	S	Std
ICV12 - 1	ICV12	21797	3.8246	-	7471B	PPB	30 Aug 2022 10:25:50	U	SMPL
ICB12 - 1	ICB12	-54	0.0083	-	7471B	PPB	30 Aug 2022 10:28:04	U	SMPL
CCV25 - 1	CCV25	28789	5.0458	-	7471B	PPB	30 Aug 2022 10:30:21	U	SMPL
CCB25 - 1	CCB25	-63	0.0067	-	7471B	PPB	30 Aug 2022 10:32:36	U	SMPL
CRA - 1	CRA	900	0.1749	-	7471B	PPB	30 Aug 2022 10:34:53	U	SMPL
HighStd - 1	HighStd	57496	10.0596	-	7471B	PPB	30 Aug 2022 10:37:08	U	SMPL
ChkStd - 1	ChkStd	40734	7.1321	-	7471B	PPB	30 Aug 2022 10:39:24	U	SMPL
PB147350BL - 1	PBS	-136	-0.0060	-	7471B	PPB	30 Aug 2022 10:41:42	U	SMPL
PB147350BS - 1	LCSS	25849	4.5323	-	7471B	PPB	30 Aug 2022 10:43:59	U	SMPL
N4379-01 - 1	SB-01-1.5-2.5	1403	0.2628	-	7471B	PPB	30 Aug 2022 10:46:16	U	SMPL
N4379-03 - 1	SB-02-1-2	16408	2.8834	-	7471B	PPB	30 Aug 2022 10:48:34	U	SMPL
N4379-05 - 1	SB-03-9-11	361	0.0808	-	7471B	PPB	30 Aug 2022 10:50:51	U	SMPL
N4379-05DUP - 1	SB-03-9-11DUP	180	0.0492	-	7471B	PPB	30 Aug 2022 10:53:09	U	SMPL
N4379-06 - 1	N4379-05MS	21312	3.7399	-	7471B	PPB	30 Aug 2022 10:55:25	U	SMPL
CCV26 - 1	CCV26	28855	5.0573	-	7471B	PPB	30 Aug 2022 10:57:40	U	SMPL
CCB26 - 1	CCB26	-64	0.0065	-	7471B	PPB	30 Aug 2022 10:59:57	U	SMPL
N4379-07 - 1	N4379-05MSD	23510	4.1238	-	7471B	PPB	30 Aug 2022 11:02:13	U	SMPL
N4379-08 - 1	SB-04-12-13	739	0.1468	-	7471B	PPB	30 Aug 2022 11:04:28	U	SMPL
N4379-09 - 1	SB-05-2-4	1799	0.3319	-	7471B	PPB	30 Aug 2022 11:06:45	U	SMPL
N4379-11 - 1	DUP-X	722	0.1438	-	7471B	PPB	30 Aug 2022 11:09:01	U	SMPL
N4379-12 - 1	SB-06-9.5-11	192	0.0513	-	7471B	PPB	30 Aug 2022 11:11:16	U	SMPL
N4379-13 - 1	SB-07-3-5	53700	9.3966	-	7471B	PPB	30 Aug 2022 11:13:33	U	SMPL
N4379-15 - 1	SB-08-9-11	432	0.0932	-	7471B	PPB	30 Aug 2022 11:15:50	U	SMPL
N4379-16 - 1	SB-09-1-2	8605	1.5206	-	7471B	PPB	30 Aug 2022 11:18:08	U	SMPL
N4379-18 - 1	SB-10-1.5-2.5	44358	7.7650	-	7471B	PPB	30 Aug 2022 11:22:25	U	SMPL
N4397-02 - 1	VNJ-216	2691	0.4877	-	7471B	PPB	30 Aug 2022 11:24:40	U	SMPL
CCV27 - 1	CCV27	28954	5.0746	-	7471B	PPB	30 Aug 2022 11:32:14	U	SMPL
CCB27 - 1	CCB27	-90	0.0020	-	7471B	PPB	30 Aug 2022 11:34:29	U	SMPL
N4398-01 - 1	ROCK-FINES-QA/QC-COMP	2187	0.3997	-	7471B	PPB	30 Aug 2022 11:36:47	U	SMPL
N4414-01 - 1	OK-2-082622	4332	0.7743	-	7471B	PPB	30 Aug 2022 11:39:02	U	SMPL
N4415-01 - 1	OR-02-082622	3884	0.6961	-	7471B	PPB	30 Aug 2022 11:41:18	U	SMPL
N4416-01 - 1	B	9402	1.6598	-	7471B	PPB	30 Aug 2022 11:43:34	U	SMPL
N4417-01 - 1	F	3076	0.5550	-	7471B	PPB	30 Aug 2022 11:45:49	U	SMPL
N4379-05LX5 - 1		-129	-0.0048	-	7471B	PPB	30 Aug 2022 11:48:06	U	SMPL
N4379-05A - 1		22289	3.9106	-	7471B	PPB	30 Aug 2022 11:50:23	U	SMPL
CCV28 - 1	CCV28	28754	5.0397	-	7471B	PPB	30 Aug 2022 11:52:39	U	SMPL
CCB28 - 1	CCB28	-50	0.0090	-	7471B	PPB	30 Aug 2022 11:54:57	U	SMPL