



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121789

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	08/30/2022	09:10
2	N4427-01		1	12	100.00	>212.0	>212.0	08/30/2022	09:30
3	N4437-01		1	13	100.00	>212.0	>212.0	08/30/2022	10:05
4	N4437-01DUP		1	12	100.00	>212.0	>212.0	08/30/2022	10:35
5	N4437-02		1	14	100.00	>212.0	>212.0	08/30/2022	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 121785

WorkList Name : FLASH-WATER-083022

WorkList ID : 162767

Department : Wet-Chemistry

Date : 08-30-2022 09:00:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/05/2022	Water	N4427-01	Flash Point	Cool 4 deg C	PSEG03	M11	D-171	08/29/2022	1010B
09/12/2022	Water	N4437-01	Flash Point	Cool 4 deg C	TECT01	M11	MW-2-(17)	08/29/2022	1010B
09/12/2022	Water	N4437-02	Flash Point	Cool 4 deg C	TECT01	M11	BASMENT-STAND-PIPE	08/29/2022	1010B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

8/30/2022 9:05
SS (Wetchem)
GV (SM)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

8/30/2022 11:30
GV (SM)
SJ (Wetchem)