

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB121791

ANALYST: Iwona

SUPERVISOR REVIEW BY: Jignesh

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2843
IODINE SOLUTION .025N 1L	W2924
Starch Solution, 4L	W2766

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB147360BL		1	50	50	2.00	0.00	1.96	1.96	0.04	0.00	0.00	08/30/2022	11:20
2	N4371-01		1	50	50	2.00	1.96	3.88	1.92	0.08	0.04	0.32	08/30/2022	11:23
3	N4371-01DUP		1	50	50	2.00	3.88	5.82	1.94	0.06	0.02	0.16	08/30/2022	11:26
4	N4427-01		1	50	50	2.00	5.88	7.70	1.82	0.18	0.14	1.12	08/30/2022	11:29

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume