



TOTAL SOLIDS - SM2540B

SUPERVISOR: Sohil

ANALYST: jignesh

Date: 08/30/2022

Run Number: LB121792

BalanceID: WC SC-6

OvenID: WC OVEN-1

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 08/30/2022 12:00 TEMP1 OUT: 103 °C 08/30/2022 13:00
TEMP2 IN: 103 °C 08/30/2022 13:30 TEMP2 OUT: 103 °C 08/30/2022 14:30
TEMP3 IN: 104 °C 08/30/2022 16:20 TEMP3 OUT: 103 °C 08/31/2022 08:00
TEMP4 IN: 104 °C 08/31/2022 08:30 TEMP4 OUT: 104 °C 08/31/2022 10:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB121792BL	LB121792BL	91.4563	91.4563	100	91.4563	91.4563	91.4563	0.0000	0
2	N4437-01	MW-2-(17)	89.9898	89.9898	100	90.0429	90.0429	90.0429	0.0531	531
3	N4437-01DUP	MW-2-(17) DUP	85.0765	85.0765	100	85.1303	85.1303	85.1303	0.0538	538
4	N4437-02	BASMENT-STAND-PIPE	89.9971	89.9971	100	90.0957	90.0957	90.0957	0.0986	986

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

121792

WorkList Name : ts n4437

WorkList ID : 162769

Department : Wet-Chemistry

Date : 08-30-2022 11:28:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/05/2022	Water	N4437-01	TS	Cool 4 deg C	TECT01	M11	MW-2-(17)	08/29/2022	SM2540 B
09/05/2022	Water	N4437-02	TS	Cool 4 deg C	TECT01	M11	BASMENT-STAND-PIPE	08/29/2022	SM2540 B

Date/Time 08/30/22 12:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 08/30/22 16:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]