



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB121798

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP98747
Turbidity Calibration std, 0NTU	WP98746
Turbidity Calibration std, 20NTU	WP98741
Turbidity Calibration std, 40NTU	WP98740
Turbidity Calibration std, 1NTU	WP98744
Turbidity Calibration std, 5NTU	WP98743
Turbidity Calibration - CCV std, 10 NTU	WP98742
10 NTU Standard 500 ml	W2752

Intercept: 0.0483

Slope: 0.9323

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.11		08/30/2022	12:40
2	CAL2	1	1	1.130	1.16	-16	08/30/2022	12:43
3	CAL3	5	1	4.410	4.68	6.4	08/30/2022	12:46
4	CAL4	10	1	9.460	10.10	-1	08/30/2022	12:50
5	CAL5	20	1	18.600	19.90	0.5	08/30/2022	12:53
6	CAL6	40	1	37.400	40.06	-0.2	08/30/2022	12:56



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB121798

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.360	9.988	08/30/2022	13:00
2	ICB		1	0.142	0.101	08/30/2022	13:03
3	CCV1	10	1	9.380	10.009	08/30/2022	13:06
4	CCB1		1	0.148	0.107	08/30/2022	13:10
5	LB121798BL		1	0.150	0.109	08/30/2022	13:13
6	N3980-12		1	0.528	0.515	08/30/2022	13:16
7	CCV2	10	1	9.240	9.859	08/30/2022	13:20
8	CCB2		1	0.144	0.103	08/30/2022	13:23