



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Sohil

Run Number: LB121803

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	08/30/2022	13:40
2	CAL2	1	Water	NA	NA	20.8	7.00	08/30/2022	13:41
3	CAL3	1	Water	NA	NA	20.8	10.03	08/30/2022	13:42
4	ICV	1	Water	NA	NA	20.7	6.99	08/30/2022	13:45
5	CCV1	1	Water	NA	NA	20.7	2.01	08/30/2022	13:46
6	N4438-01	1	Solid	20.02	20	23.5	6.90	08/30/2022	13:50
7	N4438-01DUP	1	Solid	20.03	20	23.6	6.91	08/30/2022	13:51
8	N4438-02	1	Solid	20.04	20	23.4	7.42	08/30/2022	14:00
9	N4438-03	1	Solid	20.03	20	23.6	7.89	08/30/2022	14:05
10	CCV2	1	Water	NA	NA	20.7	12.03	08/30/2022	14:10

VB 121803

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N4438

WorkList ID : 162776

Department : Wet-Chemistry

Date : 08-30-2022 13:29:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/29/2022	Solid	N4438-01	Corrosivity	Cool 4 deg C	PARS15	M11	KTS2-WC-S-TR-01-02-03	08/29/2022	9045D
08/29/2022	Solid	N4438-02	Corrosivity	Cool 4 deg C	PARS15	M11	KTS2-WC-S-MH-01-0-20	08/29/2022	9045D
08/29/2022	Solid	N4438-03	Corrosivity	Cool 4 deg C	PARS15	M11	KTS2-WC-S-MH-02-0-20	08/29/2022	9045D

Date/Time 08/30/22 13:35
Raw Sample Received by: JFCSM
Raw Sample Relinquished by: JFCSM

Date/Time 08/30/22 17:00
Raw Sample Received by: JFCSM
Raw Sample Relinquished by: JFCSM