

LB121818


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43187



A= 0.0000e+000

B= 2.2755e-004

C= 2.8454e-002

Rho= 0.9995512

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.017	-0.017	-201	0.000	-201					—
0.2	0.200	0.228	0.028	875	0.0 %	875					14
2.5	2.500	2.440	-0.060	10596	0.0 %	10596					-2
5.0	5.000	4.977	-0.023	21748	0.0 %	21748					0
7.5	7.500	7.717	0.217	33790	0.0 %	33790					3
10.0	10.000	9.856	-0.144	43187	0.0 %	43187					-1

LB121818 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Aug 2022 11:32:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-201	-	0.0000	7470A	PPB	31 Aug 2022 13:31:41	S	Std
0.2 - 1	50.2	875	-	0.2000	7470A	PPB	31 Aug 2022 13:34:01	S	Std
2.5 - 1	52.5	10596	-	2.5000	7470A	PPB	31 Aug 2022 13:36:18	S	Std
5.0 - 1	55	21748	-	5.0000	7470A	PPB	31 Aug 2022 13:38:35	S	Std
7.5 - 1	57.5	33790	-	7.5000	7470A	PPB	31 Aug 2022 13:40:51	S	Std
10.0 - 1	511	43187	-	10.0000	7470A	PPB	31 Aug 2022 13:52:15	S	Std
ICV13 - 1	ICV13	16978	3.8918	-	7470A	PPB	31 Aug 2022 13:55:00	U	SMPL
ICB13 - 1	ICB13	-211	-0.0196	-	7470A	PPB	31 Aug 2022 13:57:14	U	SMPL
CCV29 - 1	CCV29	22214	5.0832	-	7470A	PPB	31 Aug 2022 13:59:30	U	SMPL
CCB29 - 1	CCB29	-212	-0.0198	-	7470A	PPB	31 Aug 2022 14:01:45	U	SMPL
CRA - 1	CRA	752	0.1996	-	7470A	PPB	31 Aug 2022 14:04:02	U	SMPL
HighStd - 1	HighStd	43735	9.9803	-	7470A	PPB	31 Aug 2022 14:08:52	U	SMPL
ChkStd - 1	ChkStd	31618	7.2231	-	7470A	PPB	31 Aug 2022 14:11:08	U	SMPL
PB147391BL - 1	PBW	-243	-0.0268	-	7470A	PPB	31 Aug 2022 14:13:25	U	SMPL
PB147391BS - 1	LCSW	18710	4.2859	-	7470A	PPB	31 Aug 2022 14:15:42	U	SMPL
N4421-02 - 1	WC-PALM-SP-01	-16	0.0248	-	7470A	PPB	31 Aug 2022 14:17:59	U	SMPL
N4421-07 - 1	WC-PALM-SP-02	-61	0.0146	-	7470A	PPB	31 Aug 2022 14:20:16	U	SMPL
N4421-12 - 1	WC-PALM-SP-03	3	0.0291	-	7470A	PPB	31 Aug 2022 14:22:33	U	SMPL
N4423-02 - 1	WC-1	-84	0.0093	-	7470A	PPB	31 Aug 2022 14:24:50	U	SMPL
N4423-05 - 1	WC-2	-100	0.0057	-	7470A	PPB	31 Aug 2022 14:27:05	U	SMPL
CCV30 - 1	CCV30	21823	4.9943	-	7470A	PPB	31 Aug 2022 14:29:21	U	SMPL
CCB30 - 1	CCB30	-202	-0.0175	-	7470A	PPB	31 Aug 2022 14:31:36	U	SMPL
N4423-05DUP - 1	WC-2DUP	-102	0.0052	-	7470A	PPB	31 Aug 2022 14:33:53	U	SMPL
N4423-05MS - 1	WC-2MS	15833	3.6312	-	7470A	PPB	31 Aug 2022 14:36:09	U	SMPL
N4423-05MSD - 1	WC-2MSD	16115	3.6954	-	7470A	PPB	31 Aug 2022 14:38:25	U	SMPL
N4423-05LX5 - 1		-272	-0.0334	-	7470A	PPB	31 Aug 2022 14:40:40	U	SMPL
N4423-05A - 1		14963	3.4333	-	7470A	PPB	31 Aug 2022 14:42:57	U	SMPL
PB147339TB - 1	PB147339TB	-174	-0.0111	-	7470A	PPB	31 Aug 2022 14:45:13	U	SMPL
CCV31 - 1	CCV31	21539	4.9296	-	7470A	PPB	31 Aug 2022 14:47:29	U	SMPL
CCB31 - 1	CCB31	-208	-0.0189	-	7470A	PPB	31 Aug 2022 14:49:47	U	SMPL