



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB121822

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	84.8	08/31/2022	09:10
2	N4447-01		1	12	100.00	>212.0	>212.0	08/31/2022	09:50
3	N4447-01DUP		1	15	100.00	>212.0	>212.0	08/31/2022	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121822

WorkList Name : FLASHPOINT-08-31-22

WorkList ID : 162816

Department : Wet-Chemistry

Date : 08-31-2022 09:00:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/02/2022	Solid	N4447-01	Flash Point	Cool 4 deg C	PSEG03	M11	AUD-1656-OIL	08/30/2022	1010B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

8/30/2022 9:00
SJ (wetchem)
SN (SM)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

8/30/2022 11:30
GAL (SM)
SJ (wetchem)